

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-006688

Location: STEE  
KIDDER Co. S. O.

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 996 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 03417080A3 Cyl #: 28 Exp. Date: 3/5/19  
Test 1 0.082 Test 4 0.082 Test 7 0.081  
Test 2 0.082 Test 5 0.082 Test 8 0.081  
Test 3 0.082 Test 6 0.082 Test 9 0.081  
Average 0.082

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

[Signature]  
Breath Analyst Signature

5.3.18  
Date

[Signature]  
Reviewed by

5.9.18  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
05/03/2018      13:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:53
02 Std. Sol.	0.018	13:54
03 Room Air	0.000	13:54
04 Std. Sol.	0.018	13:55
05 Room Air	0.000	13:55
06 Std. Sol.	0.019	13:56
07 Room Air	0.000	13:57

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 22

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
05/03/2018      13:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:59
02 Std. Sol.	0.300	14:00
03 Room Air	0.000	14:00
04 Std. Sol.	0.301	14:01
05 Room Air	0.000	14:01
06 Std. Sol.	0.301	14:02
07 Room Air	0.000	14:03

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 22

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
05/03/2018      13:43

Test	AC	Time
01 Room Air	0.000	13:43
02 *Subject Test	INT*	13:44
03 Room Air	0.000	13:45

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 03417080A3  
Cyl No = 28  
Expiration Date = 03/05/2019  
County = 22      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006688  
Location = STEE                    8164.14.00 09/16  
05/03/2018                            13:33

Test	AC	Time
01 Diagnostic	OK	13:34
02 Room Air	0.000	13:35
03 *Subject Test	RFI*	13:35
04 Room Air	0.000	13:35

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 03417080A3  
Cyl No = 28  
Expiration Date = 03/05/2019  
County = 22                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
05/03/2018      13:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Gas	0.082	13:46
03 Room Air	0.000	13:46
04 Std. Gas	0.082	13:47
05 Room Air	0.000	13:47
06 Std. Gas	0.082	13:47
07 Room Air	0.000	13:48

Lot No = 03417080A3  
Cyl No = 28  
Exp Date = 03/05/2019  
County = 22      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
05/03/2018      13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:49
02 Std. Gas	0.082	13:49
03 Room Air	0.000	13:50
04 Std. Gas	0.082	13:50
05 Room Air	0.000	13:51
06 Std. Gas	0.082	13:51
07 Room Air	0.000	13:51

Lot No = 03417080A3  
Cyl No = 28  
Exp Date = 03/05/2019  
County = 22      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
05/03/2018      14:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:04
02 Std. Gas	0.081	14:04
03 Room Air	0.000	14:05
04 Std. Gas	0.081	14:05
05 Room Air	0.000	14:06
06 Std. Gas	0.081	14:06
07 Room Air	0.000	14:06

Lot No = 03417080A3  
Cyl No = 28  
Exp Date = 03/05/2019  
County = 22      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.0804C

Form 106-I8000