

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006689

Location: LINT
EMMONS COUNTY
S.O.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1001 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 1 Exp. Date: 9/5/19

Test 1	<u>0.081</u>	Test 4	<u>0.082</u>	Test 7	<u>0.082</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.082</u>
Test 3	<u>0.082</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
Average <u>0.082</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edle
Breath Analyst Signature

5.3.2018
Date

Roberta Luque-Tummo
Reviewed by

5.9.2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19817080A4
Standard Cyl #? 1
Standard Expiration? 09/05/2019
Oper No? 666666

Flow Cal. Date: 09/19/2017
Slope 625
Intercept -591019

IR Calibration Date: 09/19/2017
 3um 9um

0th Coef(*100): -14072 -18220
1st Coef(*100): 273930 133081
2nd Coef(*100): 2533 1211
H2O adj(mg/l*10k): 588 599

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:42
02 Std. Sol.	0.020	10:43
03 Room Air	0.000	10:44
04 Std. Sol.	0.020	10:44
05 Room Air	0.000	10:45
06 Std. Sol.	0.020	10:46
07 Room Air	0.000	10:46

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Sol.	0.305	10:48
03 Room Air	0.000	10:49
04 Std. Sol.	0.305	10:49
05 Room Air	0.000	10:50
06 Std. Sol.	0.306	10:51
07 Room Air	0.000	10:51

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:37

Test	AC	Time
01 Room Air	0.000	10:38
02 *Subject Test	INT*	10:39
03 Room Air	0.000	10:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 1
Expiration Date = 09/05/2019
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:34

Test	AC	Time
01 Diagnostic	OK	10:35
02 Room Air	0.000	10:35
03 *Subject Test	RFI*	10:35
04 Room Air	0.000	10:36

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 1
Expiration Date = 09/05/2019
County = 15 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:53
02 Std. Gas	0.081	10:53
03 Room Air	0.000	10:54
04 Std. Gas	0.082	10:54
05 Room Air	0.000	10:55
06 Std. Gas	0.082	10:55
07 Room Air	0.000	10:55

Lot No = 19817080A4
Cyl No = 1
Exp Date = 09/05/2019
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:56
02 Std. Gas	0.082	10:57
03 Room Air	0.000	10:57
04 Std. Gas	0.081	10:58
05 Room Air	0.000	10:58
06 Std. Gas	0.082	10:59
07 Room Air	0.000	10:59

Lot No = 19817080A4
Cyl No = 1
Exp Date = 09/05/2019
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = LINT 8164.14.00 09/16
05/03/2018 10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.082	11:00
03 Room Air	0.000	11:01
04 Std. Gas	0.082	11:01
05 Room Air	0.000	11:02
06 Std. Gas	0.082	11:02
07 Room Air	0.000	11:03

Lot No = 19817080A4
Cyl No = 1
Exp Date = 09/05/2019
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.086 AC

Form 106-I8000