

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006691

Location: STAT
MERCER County
S.O.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1000 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 03417080A3 Cyl #: 7 Exp. Date: 3/5/2019

Test 1	<u>0.081</u>	Test 4	<u>0.082</u>	Test 7	<u>0.082</u>
Test 2	<u>0.081</u>	Test 5	<u>0.082</u>	Test 8	<u>0.082</u>
Test 3	<u>0.081</u>	Test 6	<u>0.082</u>	Test 9	<u>0.082</u>
Average <u>0.082</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eden
Breath Analyst Signature

5-4-18
Date

Roberta Huggins-Nummo
Reviewed by

5-9-18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 12:46

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 007
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 09/07/2016
Slope 715
Intercept -719337

IR Calibration Date: 09/07/2016

	3um	9um
0th Coef(*100):	-11169	-23079
1st Coef(*100):	265834	136373
2nd Coef(*100):	1816	1210
H2O adj(mg/l*10k):	497	550

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 13:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Sol.	0.017	13:12
03 Room Air	0.000	13:12
04 Std. Sol.	0.017	13:13
05 Room Air	0.000	13:14
06 Std. Sol.	0.018	13:14
07 Room Air	0.000	13:15

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 13:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:06
02 Std. Sol.	0.292	13:07
03 Room Air	0.000	13:08
04 Std. Sol.	0.294	13:08
05 Room Air	0.000	13:09
06 Std. Sol.	0.295	13:10
07 Room Air	0.000	13:10

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 12:49

Test	AC	Time
01 Room Air	0.000	12:49
02 *Subject Test	INT*	12:50
03 Room Air	0.000	12:51

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 007
Expiration Date = 03/05/2019
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 12:46

Test	AC	Time
01 Diagnostic	OK	12:47
02 Room Air	0.000	12:48
03 *Subject Test	RFI*	12:48
04 Room Air	0.000	12:48

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 007
Expiration Date = 03/05/2019
County = 29 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 12:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:52
02 Std. Gas	0.081	12:53
03 Room Air	0.000	12:53
04 Std. Gas	0.081	12:53
05 Room Air	0.000	12:54
06 Std. Gas	0.081	12:54
07 Room Air	0.000	12:55

Lot No = 03417080A3
Cyl No = 007
Exp Date = 03/05/2019
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 13:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:16
02 Std. Gas	0.082	13:16
03 Room Air	0.000	13:17
04 Std. Gas	0.082	13:17
05 Room Air	0.000	13:18
06 Std. Gas	0.082	13:18
07 Room Air	0.000	13:19

Lot No = 03417080A3
Cyl No = 7
Exp Date = 03/05/2019
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = STAT 8164.14.00 09/16
05/04/2018 13:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Gas	0.082	13:20
03 Room Air	0.000	13:20
04 Std. Gas	0.082	13:21
05 Room Air	0.000	13:21
06 Std. Gas	0.082	13:22
07 Room Air	0.000	13:22

Lot No = 03417080A3
Cyl No = 7
Exp Date = 03/05/2019
County = 29 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000