

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006511

Location: ROLL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 963 psi Regulator: 950 psi

ROLLETTE COUNTY
CEE S.O.

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.360 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 0341708043 Cyl #: 12 Exp. Date: 3.5.2019

Test 1 0.082 Test 4 0.081 Test 7 0.082

Test 2 0.081 Test 5 0.081 Test 8 0.082

Test 3 0.081 Test 6 0.082 Test 9 0.082

Average 0.082

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eds
Breath Analyst Signature

5-9-2018
Date

Roberta Muger-Nunno
Reviewed By

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:00

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 12
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 08/21/2015
Slope 688
Intercept -714727

IR Calibration Date: 08/21/2015
 3um 9um

0th Coef(*100): -19744 -15946
1st Coef(*100): 272125 142950
2nd Coef(*100): 2975 1383
H2O adj(mg/l*10k): 725 394

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Sol.	0.017	13:24
03 Room Air	0.000	13:25
04 Std. Sol.	0.018	13:26
05 Room Air	0.000	13:26
06 Std. Sol.	0.018	13:27
07 Room Air	0.000	13:27


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 40

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Sol.	0.302	13:19
03 Room Air	0.000	13:20
04 Std. Sol.	0.305	13:21
05 Room Air	0.000	13:21
06 Std. Sol.	0.305	13:22
07 Room Air	0.000	13:23

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 40

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:08

Test	AC	Time
01 Room Air	0.000	13:08
02 *Subject Test	INT*	13:09
03 Room Air	0.000	13:10

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 03417080A3
Cyl No = 12
Expiration Date = 03/05/2019
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 Ac + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:04

Test	AC	Time
01 Diagnostic	OK	13:06
02 Room Air	0.000	13:06
03 *Subject Test	RFI*	13:06
04 Room Air	0.000	13:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 03417080A3
Cyl No = 12
Expiration Date = 03/05/2019
County = 40 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:10
02 Std. Gas	0.082	13:11
03 Room Air	0.000	13:11
04 Std. Gas	0.081	13:12
05 Room Air	0.000	13:12
06 Std. Gas	0.081	13:12
07 Room Air	0.000	13:13

Lot No = 03417080A3
Cyl No = 12
Exp Date = 03/05/2019
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:14
02 Std. Gas	0.081	13:14
03 Room Air	0.000	13:15
04 Std. Gas	0.081	13:15
05 Room Air	0.000	13:16
06 Std. Gas	0.082	13:16
07 Room Air	0.000	13:17

Lot No = 03417080A3
Cyl No = 12
Exp Date = 03/05/2019
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = ROLL 8164.14.00 09/16
05/09/2018 13:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.082	13:29
03 Room Air	0.000	13:30
04 Std. Gas	0.082	13:30
05 Room Air	0.000	13:31
06 Std. Gas	0.082	13:31
07 Room Air	0.000	13:31

Lot No = 03417080A3
Cyl No = 12
Exp Date = 03/05/2019
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.082 AC

Form 106-I8000