

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006686

Location: GRAF
WALSH COUNTY, S.O.,
L.I.C.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1008 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 22 Exp. Date: 9.5.2019

Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.083</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eda
Breath Analyst Signature

5-10-2018
Date

Roberta Hugo-Nimmo
Reviewed by

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 13:37

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19817080A4
Standard Cyl #? 22
Standard Expiration? 09/05/2019
Oper No? 666666

Flow Cal. Date: 09/07/2016
Slope 770
Intercept -749410

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100):	-14519	-24061
1st Coef(*100):	252211	131592
2nd Coef(*100):	3527	1490
H2O adj(mg/l*10k):	522	594

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 13:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Sol.	0.020	13:49
03 Room Air	0.000	13:50
04 Std. Sol.	0.019	13:50
05 Room Air	0.000	13:51
06 Std. Sol.	0.020	13:52
07 Room Air	0.000	13:52
08 Sim Temp = 34.0°C		

Simul Ser No = DR3377
Std Sol No = 201610A
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 14:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Sol.	0.295	14:04
03 Room Air	0.000	14:05
04 Std. Sol.	0.296	14:05
05 Room Air	0.000	14:06
06 Std. Sol.	0.296	14:07
07 Room Air	0.000	14:07


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 50

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 13:41

Test	AC	Time
01 Room Air	0.000	13:41
02 *Subject Test	INT*	13:42
03 Room Air	0.000	13:43

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 19817080A4
Cyl No = 22
Expiration Date = 09/05/2019
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 13:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Gas	0.082	13:44
03 Room Air	0.000	13:44
04 Std. Gas	0.082	13:45
05 Room Air	0.000	13:45
06 Std. Gas	0.082	13:46
07 Room Air	0.000	13:46

Lot No = 19817080A4
Cyl No = 22
Exp Date = 09/05/2019
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.082	14:09
03 Room Air	0.000	14:10
04 Std. Gas	0.083	14:10
05 Room Air	0.000	14:11
06 Std. Gas	0.082	14:11
07 Room Air	0.000	14:12

Lot No = 19817080A4
Cyl No = 22
Exp Date = 09/05/2019
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/10/2018 14:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Gas	0.082	14:13
03 Room Air	0.000	14:13
04 Std. Gas	0.082	14:14
05 Room Air	0.000	14:14
06 Std. Gas	0.082	14:15
07 Room Air	0.000	14:15

Lot No = 19817080A4
Cyl No = 22
Exp Date = 09/05/2019
County = 50 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000