

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006512

Location: CARR
CARRINGTON P.D.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1002 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5-15-19

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 20016080 A2 Cyl #: 44 Exp. Date: 9/5/18

Test 1	<u>0.080</u>	Test 4	<u>0.082</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.080</u>
Average <u>0.081</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Reviewed by

Date

Date

5-15-18

16 May 18

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:29

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 20016080A2
Standard Cyl #? 44
Standard Expiration? 09/05/2018
Oper No? 666666

Flow Cal. Date: 08/20/2015
Slope 680
Intercept -660856

IR Calibration Date: 08/21/2015
 3um 9um

	3um	9um
0th Coef(*100):	-23051	-19261
1st Coef(*100):	266836	137135
2nd Coef(*100):	3974	1466
H2O adj(mg/l*10k):	785	487

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Sol.	0.017	13:48
03 Room Air	0.000	13:49
04 Std. Sol.	0.018	13:50
05 Room Air	0.000	13:50
06 Std. Sol.	0.018	13:51
07 Room Air	0.000	13:51


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 16

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:53
02 Std. Sol.	0.291	13:54
03 Room Air	0.000	13:55
04 Std. Sol.	0.294	13:56
05 Room Air	0.000	13:56
06 Std. Sol.	0.295	13:57
07 Room Air	0.000	13:58


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 16

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:34

Test	AC	Time
01 Room Air	0.000	13:35
02 *Subject Test	INT*	13:36
03 Room Air	0.000	13:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 20016080A2
Cyl No = 44
Expiration Date = 09/05/2018
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:30

Test	AC	Time
01 Diagnostic	OK	13:31
02 Room Air	0.000	13:32
03 *Subject Test	RFI*	13:32
04 Room Air	0.000	13:32

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 20016080A2
Cyl No = 44
Expiration Date = 09/05/2018
County = 16 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Gas	0.080	13:39
03 Room Air	0.000	13:40
04 Std. Gas	0.081	13:40
05 Room Air	0.000	13:41
06 Std. Gas	0.081	13:41
07 Room Air	0.000	13:42

Lot No = 20016080A2
Cyl No = 44
Exp Date = 09/05/2018
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Gas	0.082	13:43
03 Room Air	0.000	13:43
04 Std. Gas	0.081	13:44
05 Room Air	0.000	13:44
06 Std. Gas	0.081	13:45
07 Room Air	0.000	13:45

Lot No = 20016080A2
Cyl No = 44
Exp Date = 09/05/2018
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/15/2018 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:00
02 Std. Gas	0.081	14:00
03 Room Air	0.000	14:01
04 Std. Gas	0.081	14:01
05 Room Air	0.000	14:02
06 Std. Gas	0.080	14:02
07 Room Air	0.000	14:03

Lot No = 20016080A2

Cyl No = 44

Exp Date = 09/05/2018

County = 16

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000