

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006677

Location: WARD  
WARD COUNTY S.O.  
JAIL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 766 psi Regulator: 750 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5-15-19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 19817080A4 Cyl #: 23 Exp. Date: 9-5-19

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eder  
Breath Analyst Signature

5-17-2018  
Date

Roberta Yungue Nimmo  
Reviewed by

24 May 18  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-006677  
Location = WARD                    8164.14.00 09/16  
05/17/2018                                  14:28

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                 Yes  
Inhib Printer(Y/N)?                 No  
Display Volume?                     No  
Disable On Memfull?                 Yes  
# of Print Copies?                   1  
Select Std (D/W/I)?                 Dry  
Standard Value?                     0.080  
Standard Lot #?                      19817080A4  
Standard Cyl #?                      23  
Standard Expiration?                09/05/2019  
Oper No?                             666666

Flow Cal. Date:                     08/24/2016  
Slope                                696  
Intercept                            -655873

IR Calibration Date:                08/24/2016

	3um	9um
0th Coef(*100):	-12424	-22862
1st Coef(*100):	263434	131353
2nd Coef(*100):	5070	1477
H2O adj(mg/l*10k):	569	590

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006677  
Location = WARD                8164.14.00 09/16  
05/17/2018                        14:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Sol.	0.019	14:41
03 Room Air	0.000	14:42
04 Std. Sol.	0.019	14:43
05 Room Air	0.000	14:43
06 Std. Sol.	0.019	14:44
07 Room Air	0.000	14:44

08 Sim Temp = 34.0°C

Simul Ser No = DR3377  
Std Sol No = 201610A  
County = 51                      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *Low AC*  
                    *0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006677  
Location = WARD      8164.14.00 09/16  
05/17/2018      14:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:46
02 Std. Sol.	0.309	14:46
03 Room Air	0.000	14:47
04 Std. Sol.	0.305	14:48
05 Room Air	0.000	14:48
06 Std. Sol.	0.304	14:49
07 Room Air	0.000	14:49

08 Sim Temp = 34.0°C

Simul Ser No = DR5191  
Std Sol No = 17350  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006677  
Location = WARD      8164.14.00 09/16  
05/17/2018      14:31

Test	AC	Time
01 Room Air	0.000	14:32
02 *Subject Test	INT*	14:32
03 Room Air	0.000	14:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 23  
Expiration Date = 09/05/2019  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006677  
Location = WARD                    8164.14.00 09/16  
05/17/2018                            14:28

Test	AC	Time
01 Diagnostic	OK	14:30
02 Room Air	0.000	14:30
03 *Subject Test	RFI*	14:30
04 Room Air	0.000	14:31

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 19817080A4  
Cyl No = 23  
Expiration Date = 09/05/2019  
County = 51                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006677  
Location = WARD                    8164.14.00 09/16  
05/17/2018                            14:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:33
02 Std. Gas	0.080	14:34
03 Room Air	0.000	14:34
04 Std. Gas	0.079	14:35
05 Room Air	0.000	14:35
06 Std. Gas	0.080	14:35
07 Room Air	0.000	14:36

Lot No = 19817080A4  
Cyl No = 23  
Exp Date = 09/05/2019  
County = 51                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006677  
Location = WARD      8164.14.00 09/16  
05/17/2018      14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.080	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.080	14:38
05 Room Air	0.000	14:38
06 Std. Gas	0.080	14:39
07 Room Air	0.000	14:39

Lot No = 19817080A4  
Cyl No = 23  
Exp Date = 09/05/2019  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006677  
Location = WARD      8164.14.00 09/16  
05/17/2018      14:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:51
02 Std. Gas	0.081	14:51
03 Room Air	0.000	14:52
04 Std. Gas	0.080	14:52
05 Room Air	0.000	14:53
06 Std. Gas	0.080	14:53
07 Room Air	0.000	14:54

Lot No = 19817080A4  
Cyl No = 23  
Exp Date = 09/05/2019  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC