

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006679

Location: VALL
VALLEY CITY P. D.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1008 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 6 Exp. Date: 9/5/19

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION, PROTECTED
BREATH HOSE SLEEVE WITH TAPE (STARTING TO WEAR)

Instrument is acceptable to be used in the field.

Charles E. Ed
Breath Analyst Signature

5-18-2018
Date

Roberta Suerger - Inman
Reviewed by

24 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:37
02 Std. Sol.	0.018	12:38
03 Room Air	0.000	12:39
04 Std. Sol.	0.018	12:39
05 Room Air	0.000	12:40
06 Std. Sol.	0.018	12:41
07 Room Air	0.000	12:41


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 02

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:48
02 Std. Sol.	0.306	12:49
03 Room Air	0.000	12:50
04 Std. Sol.	0.303	12:50
05 Room Air	0.000	12:51
06 Std. Sol.	0.301	12:52
07 Room Air	0.000	12:52


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 02

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

X HIGH AC
0.300 AC
022

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:16

Test	AC	Time
01 Room Air	0.000	12:17
02 *Subject Test	INT*	12:18
03 Room Air	0.000	12:19

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERRENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 6
Expiration Date = 09/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:10

Test	AC	Time
01 Diagnostic	OK	12:11
02 Room Air	0.000	12:12
03 *Subject Test	RFI*	12:12
04 Room Air	0.000	12:12

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 6
Expiration Date = 09/05/2019
County = 02 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:21
02 Std. Gas	0.080	12:21
03 Room Air	0.000	12:22
04 Std. Gas	0.081	12:22
05 Room Air	0.000	12:22
06 Std. Gas	0.080	12:23
07 Room Air	0.000	12:23

Lot No = 19817080A4
Cyl No = 6
Exp Date = 09/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:24
02 Std. Gas	0.080	12:25
03 Room Air	0.000	12:25
04 Std. Gas	0.080	12:26
05 Room Air	0.000	12:26
06 Std. Gas	0.080	12:26
07 Room Air	0.000	12:27

Lot No = 19817080A4
Cyl No = 6
Exp Date = 09/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/18/2018 12:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:54
02 Std. Gas	0.081	12:55
03 Room Air	0.000	12:55
04 Std. Gas	0.081	12:55
05 Room Air	0.000	12:56
06 Std. Gas	0.080	12:56
07 Room Air	0.000	12:57

Lot No = 19817080A4
Cyl No = 6
Exp Date = 09/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000