

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006682

Location: BCCC
BARNES COUNTY CORRECTIONAL CENTER

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 926 psi Regulator: 925 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 2016104 Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0341708043 Cyl #: 25 Exp. Date: 3/5/19

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.080</u>	Test 9	<u>0.079</u>
Average <u>0.079</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edin
Breath Analyst Signature

5-18-2018
Date

Roberta Hunger-Nimmitt
Reviewed by

24 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:13

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 25
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 09/06/2016
Slope 701
Intercept -706193

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100): -17150 -23247
1st Coef(*100): 271726 134617
2nd Coef(*100): 831 1215
H2O adj(mg/l*10k): 393 518

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:32
02 Std. Sol.	0.017	13:33
03 Room Air	0.000	13:33
04 Std. Sol.	0.017	13:34
05 Room Air	0.000	13:34
06 Std. Sol.	0.017	13:35
07 Room Air	0.000	13:36

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Sol.	0.296	13:38
03 Room Air	0.000	13:39
04 Std. Sol.	0.296	13:40
05 Room Air	0.000	13:40
06 Std. Sol.	0.297	13:41
07 Room Air	0.000	13:41

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:16

Test	AC	Time
01 Room Air	0.000	13:17
02 *Subject Test	INT*	13:17
03 Room Air	0.000	13:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 25
Expiration Date = 03/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:14

Test	AC	Time
01 Diagnostic	OK	13:15
02 Room Air	0.000	13:15
03 *Subject Test	RFI*	13:15
04 Room Air	0.000	13:16

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 25
Expiration Date = 03/05/2019
County = 02 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Gas	0.079	13:20
03 Room Air	0.000	13:20
04 Std. Gas	0.079	13:21
05 Room Air	0.000	13:21
06 Std. Gas	0.079	13:21
07 Room Air	0.000	13:22

Lot No = 03417080A3
Cyl No = 25
Exp Date = 03/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:23
02 Std. Gas	0.080	13:23
03 Room Air	0.000	13:24
04 Std. Gas	0.079	13:24
05 Room Air	0.000	13:25
06 Std. Gas	0.080	13:25
07 Room Air	0.000	13:25

Lot No = 03417080A3
Cyl No = 25
Exp Date = 03/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
05/18/2018 13:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Gas	0.080	13:43
03 Room Air	0.000	13:44
04 Std. Gas	0.079	13:44
05 Room Air	0.000	13:45
06 Std. Gas	0.079	13:45
07 Room Air	0.000	13:45

Lot No = 03417080A3
Cyl No = 25
Exp Date = 03/05/2019
County = 02 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC