



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Patrick Renz</i>		
Location <i>BSPD</i>	Intoxilyzer® Serial Number <i>80-006675</i>	
Gas Lot Number <i>34917080A3</i>	Gas Cylinder Number <i>16</i>	Gas Expiration Date <i>2-5-2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.079* AC
 - B. *0.078* AC
 - C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Patrick Renz</i>	Date <i>5-25-18</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta M. Mims</i>	Date <i>5-25-18</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006675
Location = BSPD 8164.14.00 09/16
05/25/2018 08:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:28
02 Std. Gas	0.079	08:29
03 Room Air	0.000	08:29
04 Std. Gas	0.078	08:29
05 Room Air	0.000	08:30
06 Std. Gas	0.078	08:30
07 Room Air	0.000	08:31

Lot No = 34917080A3
Cyl No = 16
Exp Date = 02/05/2020
County = 08 Oper No. = 029710


Operator Signature
PATRICK RENZ

Remarks: *ACA New Cylinder Installed*