

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006510

Adams Co. SD  
Location: HETT

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 100 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 4 Oct 18
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 11 Oct 19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: 1CS5 Exp. Date: 15 May 19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34917080 Cyl #: 14 Exp. Date: 5 Feb 20

|                      |                     |                     |
|----------------------|---------------------|---------------------|
| Test 1 <u>0.080</u>  | Test 4 <u>0.080</u> | Test 7 <u>0.081</u> |
| Test 2 <u>0.079</u>  | Test 5 <u>0.080</u> | Test 8 <u>0.080</u> |
| Test 3 <u>0.079</u>  | Test 6 <u>0.080</u> | Test 9 <u>0.080</u> |
| Average <u>0.080</u> |                     |                     |

C. Remarks/Maintenance: Annual Inspection.

Instrument is acceptable to be used in the field.

Roberta Muegen Timmer  
Breath Analyst Signature

30 May 18  
Date

Charles E. Ede  
Reviewed by

31 MAY 18  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.14.00 09/16  
05/30/2018      14:20

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 14:21 |
| 02 Std. Sol. | 0.017 | 14:22 |
| 03 Room Air  | 0.000 | 14:22 |
| 04 Std. Sol. | 0.018 | 14:23 |
| 05 Room Air  | 0.000 | 14:23 |
| 06 Std. Sol. | 0.017 | 14:24 |
| 07 Room Air  | 0.000 | 14:25 |


08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 01

Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: LOW AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.14.00 09/16  
05/30/2018      14:11

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 14:12 |
| 02 Std. Sol. | 0.297 | 14:13 |
| 03 Room Air  | 0.000 | 14:13 |
| 04 Std. Sol. | 0.298 | 14:14 |
| 05 Room Air  | 0.000 | 14:15 |
| 06 Std. Sol. | 0.298 | 14:15 |
| 07 Room Air  | 0.000 | 14:16 |


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 01

Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: High AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006510  
Location = HETT                    8164.14.00 09/16  
05/30/2018                            14:07

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 14:07 |
| 02 *Subject Test | INT*  | 14:08 |
| 03 Room Air      | 0.000 | 14:09 |

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 14  
Expiration Date = 02/05/2020  
County = 01                            Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Interferent Check*  
*0.10 AC + 0.05% Acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006510  
Location = HETT                    8164.14.00 09/16  
05/30/2018                            14:09

| Test          | AC    | Time  |
|---------------|-------|-------|
| 01 Diagnostic | OK    | 14:10 |
| 02 Room Air   | RFI*  | 14:10 |
| 03 Room Air   | 0.000 | 14:10 |

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 14  
Expiration Date = 02/05/2020  
County = 01                            Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *RFI Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.14.00 09/16  
05/30/2018      13:55

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 13:55 |
| 02 Std. Gas | 0.080 | 13:56 |
| 03 Room Air | 0.000 | 13:56 |
| 04 Std. Gas | 0.079 | 13:57 |
| 05 Room Air | 0.000 | 13:57 |
| 06 Std. Gas | 0.079 | 13:58 |
| 07 Room Air | 0.000 | 13:58 |

Lot No = 34917080A3  
Cyl No = 14  
Exp Date = 02/05/2020  
County = 01      Oper No. = 888888

*Roberta Grieger-Nimmo*

Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.14.00 09/16  
05/30/2018      13:59

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:00 |
| 02 Std. Gas | 0.080 | 14:00 |
| 03 Room Air | 0.000 | 14:01 |
| 04 Std. Gas | 0.080 | 14:01 |
| 05 Room Air | 0.000 | 14:02 |
| 06 Std. Gas | 0.080 | 14:02 |
| 07 Room Air | 0.000 | 14:03 |

Lot No = 34917080A3  
Cyl No = 14  
Exp Date = 02/05/2020  
County = 01      Oper No. = 888888

*Roberta Grieger-Nimmo*

Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*

*0.080 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.14.00 09/16  
05/30/2018      14:03

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:03 |
| 02 Std. Gas | 0.081 | 14:04 |
| 03 Room Air | 0.000 | 14:04 |
| 04 Std. Gas | 0.080 | 14:05 |
| 05 Room Air | 0.000 | 14:05 |
| 06 Std. Gas | 0.080 | 14:06 |
| 07 Room Air | 0.000 | 14:06 |

Lot No = 34917080A3  
Cyl No = 14  
Exp Date = 02/05/2020  
County = 01      Oper No. = 888888

*Roberta Grieger-Nimmo*

Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*  
*0.080 AC*

Form 106-I8000