

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006513

Location: Hettinger Co. SO.  
MOTT

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 995 psi Regulator: 1000psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 4 Oct 18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 11 Oct 19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: 1C55 Exp. Date: 15 May 19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 20016080A Cyl #: 43 Exp. Date: 5 Sep 18

Test 1	<u>0.081</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.082</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average <u>0.081</u>					

C. Remarks/Maintenance: Annual inspection.

Instrument is acceptable to be used in the field.

Roberta Huguier-Nemms  
Breath Analyst Signature

30 May 18  
Date

Charles E. Eden  
Reviewed by

31 MAY 2018  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006513  
Location = MOTT                 8164.13.00 06/09  
05/30/2018                        11:21

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup


Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            Yes  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 20016080A2  
Standard Cyl #?                 43  
Standard Expiration?            09/05/2018  
Oper No?                        888888

Flow Cal. Date:                 08/20/2015  
Slope                            680  
Intercept                        -696037

IR Calibration Date:            08/21/2015  
                                    3um                            9um

-----  
0th Coef(\*100):                -14436                        -13610  
1st Coef(\*100):                265490                        136147  
2nd Coef(\*100):                4422                            1710  
H2O adj(mg/l\*10k):            672                             448

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = MOTT      8164.13.00 06/09  
05/30/2018      11:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Sol.	0.019	11:57
03 Room Air	0.000	11:57
04 Std. Sol.	0.018	11:58
05 Room Air	0.000	11:58
06 Std. Sol.	0.019	11:59
07 Room Air	0.000	12:00

08 Sim Temp = 34.0°C

Simul Ser No = DR3377  
Std Sol No = 201610A  
County = 21      Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = MOTT      8164.13.00 06/09  
05/30/2018      11:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Sol.	0.307	11:49
03 Room Air	0.000	11:50
04 Std. Sol.	0.306	11:51
05 Room Air	0.000	11:51
06 Std. Sol.	0.305	11:52
07 Room Air	0.000	11:53


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 21

Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: High AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = MOTT      8164.13.00 06/09  
05/30/2018      11:37

Test	AC	Time
01 Room Air	0.000	11:38
02 *Subject Test	INT*	11:39
03 Room Air	0.000	11:40

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 20016080A2  
Cyl No = 43  
Expiration Date = 09/05/2018  
County = 21      Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Interferent Check*  
*0.10 AC + 0.05% Acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006513  
Location = MOTT                    8164.13.00 06/09  
05/30/2018    11:31

Test	AC	Time
01 Diagnostic	OK	11:32
02 Room Air	0.000	11:32
03 Subject Test 1	0.000	11:33
04 Room Air	RFI*	11:34
05 Room Air	0.000	11:34

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                                    Weight = 150  
Test = DUI    Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 20016080A2  
Cyl No = 43  
Expiration Date = 09/05/2018  
County = 21    Oper No. = 888888

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: RFI Check

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006513  
Location = MOTT                 8164.13.00 06/09  
05/30/2018                        11:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Gas	0.081	11:05
03 Room Air	0.000	11:06
04 Std. Gas	0.082	11:06
05 Room Air	0.000	11:07
06 Std. Gas	0.082	11:07
07 Room Air	0.000	11:07

Lot No = 20016080A2  
Cyl No = 43  
Exp Date = 09/05/2018  
County = 21                            Oper No. = 888888

Roberta Grieger Nimmo

Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006513  
Location = MOTT                                  8164.13.00 06/09  
05/30/2018    11:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Gas	0.080	11:11
03 Room Air	0.000	11:12
04 Std. Gas	0.081	11:12
05 Room Air	0.000	11:13
06 Std. Gas	0.081	11:13
07 Room Air	0.000	11:14

Lot No = 20016080A2  
Cyl No = 43  
Exp Date = 09/05/2018  
County = 21    Oper No. = 888888

Roberta Grieger-Nimmo  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = MOTT      8164.13.00 06/09  
05/30/2018      11:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Gas	0.081	11:15
03 Room Air	0.000	11:15
04 Std. Gas	0.081	11:16
05 Room Air	0.000	11:16
06 Std. Gas	0.081	11:17
07 Room Air	0.000	11:17

Lot No = 20016080A2  
Cyl No = 43  
Exp Date = 09/05/2018  
County = 21      Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC

Form 106-I8000