

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006683

Cakes PD
Location: OK PD

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1004 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR5143 Lot #: 201610A Exp. Date: 4 Oct 18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 11 Oct 19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: 1055 Exp. Date: 15 May 19

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 19817080A4 Cyl #: 13 Exp. Date: 5 Sep 19

Test 1 0.079 Test 4 0.080 Test 7 0.079

Test 2 0.080 Test 5 0.079 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.079

Average 0.080

C. Remarks/Maintenance: Annual inspection

Instrument is acceptable to be used in the field.

Roberta Muegge-Nimmer
Breath Analyst Signature

6 Jun 18
Date

Charles E. Edr
Reviewed by

08 Jun. 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 10:52

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19817080A4
Standard Cyl #? 13
Standard Expiration? 09/05/2019
Oper No? 888888

Flow Cal. Date: 09/06/2016
Slope 694
Intercept -595721

IR Calibration Date: 09/06/2016

	3um	9um
0th Coef(*100):	-8889	-21482
1st Coef(*100):	254010	129695
2nd Coef(*100):	3922	1447
H2O adj(mg/l*10k):	608	570

***** Printer Test End *****


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 11:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Sol.	0.020	11:27
03 Room Air	0.000	11:27
04 Std. Sol.	0.019	11:28
05 Room Air	0.000	11:28
06 Std. Sol.	0.020	11:29
07 Room Air	0.000	11:29

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 201610A
County = 11 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: low AC
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 11:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Sol.	0.291	11:21
03 Room Air	0.000	11:22
04 Std. Sol.	0.292	11:22
05 Room Air	0.000	11:23
06 Std. Sol.	0.292	11:24
07 Room Air	0.000	11:24

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 11 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: High AC
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 11:05

Test	AC	Time
01 Diagnostic	OK	11:06
02 Room Air	0.000	11:06
03 *Subject Test	RFI*	11:06
04 Room Air	0.000	11:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 13
Expiration Date = 09/05/2019
County = 11 Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: RFI Check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 10:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:53
02 Std. Gas	0.079	10:54
03 Room Air	0.000	10:54
04 Std. Gas	0.080	10:54
05 Room Air	0.000	10:55
06 Std. Gas	0.080	10:55
07 Room Air	0.000	10:56

Lot No = 19817080A4
Cyl No = 13
Exp Date = 09/05/2019
County = 11 Oper No. = 888888

Roberta Grieger-Nimmo
Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.080	10:57
03 Room Air	0.000	10:58
04 Std. Gas	0.079	10:58
05 Room Air	0.000	10:58
06 Std. Gas	0.080	10:59
07 Room Air	0.000	10:59

Lot No = 19817080A4
Cyl No = 13
Exp Date = 09/05/2019
County = 11 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006683
Location = OKPD 8164.14.00 09/16
06/06/2018 10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.079	11:00
03 Room Air	0.000	11:01
04 Std. Gas	0.080	11:01
05 Room Air	0.000	11:02
06 Std. Gas	0.079	11:02
07 Room Air	0.000	11:03

Lot No = 19817080A4
Cyl No = 13
Exp Date = 09/05/2019
County = 11 Oper No. = 888888

Roberta Grieger-Nimmo

Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check

0.080 AC