

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007087

A. GENERAL Setup and Checks:

1. Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
2. Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
4. Printer Setup (Level 3,S,P)
5. Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

1. Internal Printer Test and Paper Feed (Level 1,P). Attach record.
2. External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

1. Continuous Air Blank (Level 2,A)
2. Diagnostic (Level 2,D). Attach record.
3. Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-007087 DSP Ver: 1362.14
 - Core Ver: 8164.15 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

1. DVM Monitor (Level 3,D,D): 3: 12657 s: 1 C: 47.00
9: 13118 s: 1 B: 45 0
(RFI indicator with keyed radio): 1
2. Function (F) Keys : F5 – Tone F9 – Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 952 CF: 1.06
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 173 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 47.5 I: 0.0
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 657 psi P (gauge): 650 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: MP3058 Lot #: Milli Q H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: DR7111 Lot #: 17160 Exp date: 6.8.19
Known value: 0.020 AC
Results: 0.018 AC 0.018 AC 0.018 AC
3. High AC (Level 1,C):
Sim ser #: DR3375 Lot #: 17350 Exp date: 10.11.19
Known value: 0.300 AC
Results: 0.299 AC 0.299 AC 0.299 AC

Configure dry gas standard before running tests (Level 1,S):


4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR7089 Lot #: ICS#5 Exp date: 5.15.19
Known value: 0.10AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR3846 Lot #: ACE1 Exp date: NA
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3455 Lot #: MET2 Exp date: NA
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR5189 Lot #: IPA2 Exp date: NA
Known value: 0.10% ISOPROPRANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):


Ave: 0.0796 Std Dev: 0.0005 Rel Std Dev: 0.65%

Maintenance/Repairs/Notes: INITIAL INSPECTION AFTER
RECEIVING FROM CEE NEW FROM FACTORY.

Instrument is acceptable to be used in the field.


Field Inspector's Signature

9/11/18
Date


Reviewed by

12 Sep 18
Date

Intoxilyzer Test Record and Checklist
ND00AG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 14:40

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMNPO

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNOPQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNOPQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNOPQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 80
Standard Cyl #? 1234567
Standard Expiration? 11/11/2018
Oper No? 666666

Flow Cal. Date: 07/11/2018
Slope 688
Intercept -653049

IR Calibration Date: 07/11/2018
3um 9um


0th Coef(*100): -35550 -17836

1st Coef(*100): 246580 130222

2nd Coef(*100): 4277 1696

H2O adj(mg/l*10K): 1232 531

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: INTERNAL
PRINT
TEST
Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 14:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 80
Standard Cyl #? 1234567
Standard Expiration? 11/11/2018
Oper No? 666666

Flow Cal. Date: 07/11/2018
Slope 688
Intercept -653049

IR Calibration Date: 07/11/2018

	3um	9um
0th Coef(*100):	-35550	-17836
1st Coef(*100):	246580	130222
2nd Coef(*100):	4277	1696
H2O adj(mg/l*10k):	1232	531

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

EXTERNAL PRINT
TEST


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 14:44

Test	AC	Time
01 Room Air	0.000	14:45
02 *Subject Test	REF*	14:45
03 Room Air	0.000	14:46

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = R KEY REFUSAL
Dr. Lic. = ND/DIS821456
Lot No = 80
Cyl No = 1234567
Expiration Date = 11/11/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *BAR CODE READER*
R KEY REFUSAL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 14:48

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC
CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 14:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Sol.	0.000	14:57
03 Room Air	0.000	14:58
04 Std. Sol.	0.000	14:58
05 Room Air	0.000	14:59
06 Std. Sol.	0.000	14:59
07 Room Air	0.000	15:00

08 Sim Temp = 34.0°C

Simul Ser No = MP3058
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Zero AC

0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 15:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:42
02 Std. Sol.	0.018	15:42
03 Room Air	0.000	15:43
04 Std. Sol.	0.018	15:44
05 Room Air	0.000	15:44
06 Std. Sol.	0.018	15:45
07 Room Air	0.000	15:45

08 Sim Temp = 34.0°C

Simul Ser No = DR7111
Std Sol No = 17160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 15:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:48
02 Std. Sol.	0.299	15:49
03 Room Air	0.000	15:49
04 Std. Sol.	0.299	15:50
05 Room Air	0.000	15:51
06 Std. Sol.	0.299	15:51
07 Room Air	0.000	15:52

08 Sim Temp = 34.0°C

Simul Ser No = DR3375
Std Sol No = 17350
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

H76# AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:00

Test	AC	Time
01 Room Air	0.000	16:01
02 *Subject Test	INT*	16:01
03 Room Air	0.000	16:02

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:02

Test	AC	Time
01 Room Air	0.000	16:03
02 *Subject Test	INT*	16:03
03 Room Air	0.000	16:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:05

Test	AC	Time
01 Room Air	0.000	16:05
02 *Subject Test	INT*	16:06
03 Room Air	0.000	16:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:11

Test	AC	Time
01 Room Air	0.000	16:12
02 *Subject Test	INT*	16:12
03 Room Air	0.000	16:13

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10% ISOPROPA

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:20

Test	AC	Time
01 Room Air	AMB*	16:21
02 Room Air	0.000	16:21

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

AMBIENT FAIL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:22

Test	AC	Time
01 Room Air	0.000	16:23
02 Invalid Sample	X.XXX	16:23
03 Room Air	0.000	16:24

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDLER

Remarks: *INVALID SAMPLE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:25

Test	AC	Time
01 Room Air	0.000	16:26
02 *Subject Test	ABT*	16:26
03 Room Air	0.000	16:27

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE ABORTED

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:27

Test	AC	Time
01 Room Air	0.000	16:27
02 Subject Test 1	0.021	16:28
03 Room Air	0.000	16:28
04 Reported AC	0.021	16:28

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:29

Test	AC	Time
01 Diagnostic	OK	16:29
02 Room Air	0.000	16:30
03 Subject Test 1	0.009	16:30
04 Room Air	0.000	16:32
05 Std. Gas	0.089*	16:33
06 Room Air	0.000	16:34

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL CHK OUT TOL
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK OUT
OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:35

Test	AC	Time
01 Diagnostic	OK	16:36
02 Room Air	0.000	16:37
03 *Subject Test	RFI*	16:37
04 Room Air	0.000	16:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/10/2018 16:38

Test	AC	Time
01 Diagnostic	OK	16:39
02 Room Air	0.000	16:39
03 *Subject Test	0.000*	16:42
04 Room Air	0.000	16:44
05 Std. Gas	0.080	16:45
06 Room Air	0.000	16:47
07 *Subject Test	0.000*	16:50
08 Room Air	0.000	16:50

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks: DEFICIENT
SAMPLES

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/11/2018 12:59

Test	AC	Time
01 Diagnostic	OK	13:00
02 Room Air	0.000	13:00
03 Subject Test 1	0.249*	13:00
04 Room Air	0.000	13:03
05 Std. Gas	0.080	13:04
06 Room Air	0.000	13:05
07 Subject Test 2	0.295*	13:06
08 Room Air	0.000	13:07

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: DIFFERENCE TOO
 GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/11/2018 13:09


Test	AC	Time
01 Diagnostic	OK	13:10
02 Room Air	0.000	13:10
03 Subject Test 1	0.245	13:11
04 Room Air	0.000	13:13
05 Std. Gas	0.078	13:14
06 Room Air	0.000	13:16
07 Subject Test 2	0.244	13:16
08 Room Air	0.000	13:17
09 Reported AC	0.244	13:16

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CMS TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007087
Location = TOXL 8164.15.00 08/18
09/11/2018 13:19

Test	AC	Time
01 Room Air	0.000	13:20
02 Std. Gas	0.080	13:20
03 Room Air	0.000	13:21
04 Std. Gas	0.080	13:21
05 Room Air	0.000	13:22
06 Std. Gas	0.079	13:22
07 Room Air	0.000	13:22
08 Std. Gas	0.079	13:23
09 Room Air	0.000	13:23
10 Std. Gas	0.079	13:24
11 Room Air	0.000	13:24
12 Std. Gas	0.079	13:25
13 Room Air	0.000	13:25
14 Std. Gas	0.079	13:25
15 Room Air	0.000	13:26
16 Std. Gas	0.079	13:26
17 Room Air	0.000	13:27
18 Std. Gas	0.079	13:27
19 Room Air	0.000	13:28
20 Std. Gas	0.079	13:28
21 Room Air	0.000	13:29
22 Std. Gas	0.080	13:29
23 Room Air	0.000	13:29
24 Std. Gas	0.079	13:30
25 Room Air	0.000	13:30
26 Std. Gas	0.080	13:31
27 Room Air	0.000	13:31
28 Std. Gas	0.080	13:32
29 Room Air	0.000	13:32
30 Std. Gas	0.080	13:32
31 Room Air	0.000	13:33
32 Std. Gas	0.080	13:33
33 Room Air	0.000	13:34
34 Std. Gas	0.080	13:34
35 Room Air	0.000	13:35
36 Std. Gas	0.080	13:35
37 Room Air	0.000	13:36
38 Std. Gas	0.080	13:36
39 Room Air	0.000	13:37
40 Std. Gas	0.080	13:37
41 Room Air	0.000	13:37
42 Cal Check Stats:		
43 Average	0.0796	
44 Std Dev	0.0005	
45 Rel Std Dev	0.65%	



Operator Signature
CHARLES EDER

Remarks: 0.080 AC
Lot: 19817080AH
CYL: 21

Form 106-I8000

Exp: 9-5-2019