

**CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT**

Instrument: 8000 Serial Number: 80-0070816

Agency being assigned instrument: ND OAG - Crime Lab

Date 22 Oct 18

Transferred from:
Name (print) CMI, Inc.
Signature _____
Agency CMI, Inc.

Transferred to:
Name (print) Roberta Griego Nimmo
Signature Roberta Griego Nimmo
Agency ND-OAG-Crime Lab

Reason for transfer/remarks:
- Software update was performed at CMI, Inc.

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____



SERVICE WORK ORDER

DATE	10-9-18
RCVD VIA	LHS

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

4102792

BILL TO:		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO. 585013		
PHONE		

SHIP TO: ND Crime Lab		
ATTN: Charlie Eder		
2461 E. Main Ave.		
CITY	STATE	ZIP
Bismarck	ND	58501
PHONE		
FAX		

MODEL # 8000	SERIAL # 80-507086	002480ND	BILL CODE 3
EXTRA PARTS RCVD BT, new box, Battery			
DESCRIPTION OF PROBLEM SERVICE			

Infrared / Fuel Cell PRELIMINARY TEST INFORMATION	BATTERY- EXT <u>NA</u> ; INT <u>NA</u> V
Update software. Read low on Final testing.	VERIFIED PROBLEM? Yes

WORK PERFORMED: Reflashed Mylox printer software. Installed 8164.16. Autocal and Final performed.

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	03816416	Software

Miscellaneous PRELIMINARY TEST INFORMATION
WORK PERFORMED: BT

REPAIR GTW	CAL. ADJ. JCP	CALIBRATION JCP	FINAL JCP
1/2 HOURS 2	HOURS 0.5	HOURS 1.0	HOURS 1.0
DATE Oct 15, 18	DATE Oct 16, 18	DATE Oct 16, 18	DATE Oct 16, 18

SHIPPING STICKER HERE
SPECIAL SHIPPING INSTRUCTIONS