

CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-007087

Agency being assigned instrument: ND OAG - Crime Lab

Date 22 Oct 18

Transferred from:

Transferred to:

Name (print) CMI, Inc.

Name (print) Aberta Grieger-Nimmo

Signature \_\_\_\_\_

Signature Aberta Grieger-Nimmo

Agency CMI, Inc.

Agency ND OAG - Crime Lab

Reason for transfer/remarks:

- Software update was performed at CMI, Inc.

Or Shipped Via:

USPS \_\_\_\_\_ UPS  FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_



# SERVICE WORK ORDER

DATE	10-9-12
RCVD VIA	WWS

316 E. 9<sup>th</sup> St., Owensboro KY 42303  
 Phone 866-835-0690  
 Fax: 270-685-6268

402794
--------

<b>BILL TO:</b>
ATTN:
CITY STATE ZIP
CUSTOMER NO. 585013
PHONE

<b>SHIP TO:</b> ND Crime Lab
ATTN: Charlie Edor
2461 E. Main Ave.
CITY STATE ZIP Bismarck ND 58501
PHONE
FAX

MODEL # 8000	SERIAL # 80-007087	002480 ND	BILL CODE 3
--------------	--------------------	-----------	-------------

EXTRA PARTS RCVD: RT, gas box, battery

DESCRIPTION OF PROBLEM: Service

**Infrared / Fuel Cell**  
 PRELIMINARY TEST INFORMATION  
 Update software.  
 BATTERY- EXT NA V; INT NA V  
 VERIFIED PROBLEM? Yes

WORK PERFORMED:  
 Reflashed Mylox printer software.  
 Installed 8164.16.  
 Final Only performed.

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	03816416	Software

**Miscellaneous**  
 PRELIMINARY TEST INFORMATION  
 WORK PERFORMED:  
 RT

REPAIR GTW	CAL. ADJ. ---	CALIBRATION SOP	FINAL SOP
1/2 HOURS 2	HOURS ---	HOURS 10	HOURS 10
DATE 08/12/12	DATE ---	DATE 08/16/12	DATE 08/16/12

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS