

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006507

Agency being assigned instrument: NDOAG - Crime Lab

Date 1 Nov 18

Transferred from:

Transferred to:

Name (print) Bryan Nagel

Name (print) CHARLES E. EDER

Signature *Bryan Nagel*

Signature *Charles E. Eder*

Agency McKenzie County Sheriff's Office

Agency NDOAG - Crime Lab

Reason for transfer/remarks:

80-006507 is being returned for repair.
Temp. Reg. Test failed during diagnostic.
Replacing with 80-005945

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____