



# ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Alex Cogdill</i>		
Location <i>CARRINGTON PD</i>	Intoxilyzer® Serial Number <i>80-006512</i>	
Gas Lot Number <i>34917080A3</i>	Gas Cylinder Number <i>023</i>	Gas Expiration Date <i>02/09/2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.080* AC
  - B. *0.080* AC
  - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Alex Cogdill</i>	Date <i>11/02/18</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Huerfano</i>	Date <i>11/02/18</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006512  
Location = CARR      8164.14.00 09/16  
11/02/2018      14:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.080	14:55
03 Room Air	0.000	14:55
04 Std. Gas	0.080	14:56
05 Room Air	0.000	14:56
06 Std. Gas	0.080	14:56
07 Room Air	0.000	14:57

Lot No = 34917080A3

Cyl No = 23

Exp Date = 02/05/2020

County = 16

Oper No. = 131892



Operator Signature  
ALEX COGDILL

Remarks:

Form 106-I8000