



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Allen, Kathryn Marie</i>		
Location <i>Cass Jail 45034<sup>th</sup> St So Fargo</i>	Intoxilyzer® Serial Number <i>80-006668</i>	
Gas Lot Number <i>13518080 AL0</i>	Gas Cylinder Number <i>13</i>	Gas Expiration Date <i>8-5-2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. 0.083 AC  
 B. 0.082 AC  
 C. 0.083 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn M. Allen #31087</i>	Date <i>12-3-18</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Mungo numms</i>	Date <i>12-4-18</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006668  
Location = CASS      8164.14.00 09/16  
12/03/2018      16:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:42
02 Std. Gas	0.083	16:42
03 Room Air	0.000	16:43
04 Std. Gas	0.082	16:43
05 Room Air	0.000	16:44
06 Std. Gas	0.083	16:44
07 Room Air	0.000	16:45

Lot No = 13518080A6  
Cyl No = 13  
Exp Date = 08/05/2020  
County = 09

Oper No. = 090801

  
\_\_\_\_\_  
Operator Signature  
KATHRYN ALLEN

Remarks:  
New Calibrator Installed

Form 106-I8000