



**PROPERTY DISPOSAL**  
OFFICE OF ATTORNEY GENERAL  
SFN 60567 (7/2016)

Multiple quantities of like items may be listed on one form and attach a sheet with tag numbers and serial numbers if available. Use a separate form for different items.

Division Crime Laboratory	Location of Property Breath Alcohol Program	Date 1/24/19
Point of Contact Roberta Grieger-Nimmo	Point of Contact Email rgrieger@nd.gov	Point of Contact Telephone Number 328-6159

Description of Property (make, model, color, etc.) Intoxilyzer 8000 Breath Alcohol Analyzer CMI Inc., Owensboro, KY		Tag Number (if available) AG 3566
		Serial Number (if available) 80-006668
		Date Purchased (if available) 2016
		Original Cost (if available)
		Reason for disposal Instrument displayed "EEPROM fault, EEPROM data Invalid, System Fault ID#018. It is not cost effective to repair this instrument once this error appears. Usable parts will be salvaged to repair other instruments.

Method of disposal:

- Landfill
- Electronics disposal*
- Trash can/dumpster

Signature of person disposing property <i>Roberta Grieger-Nimmo</i>
Date 24 Jan 19

*Sent to Michelle Metzger  
24 Jan 19 RGN*

**Send form to Michelle Metzger, Finance Division**

Authorized Representative Signature
Date