



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>RAY WEBER</i>		
Location <i>TRAILL COUNTY S.O. (HILL)</i>	Intoxilyzer® Serial Number <i>80-006678</i>	
Gas Lot Number <i>13518080A6</i>	Gas Cylinder Number <i>20</i>	Gas Expiration Date <i>08-05-2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.084* AC
  - B. *0.083* AC
  - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Ray Weber</i>	Date <i>03-04-2019</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Edr</i>	Date <i>03/07/2019</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006678  
Location = HILL      8164.14.00 09/16  
03/04/2019      18:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:42
02 Std. Gas	0.084	18:42
03 Room Air	0.000	18:43
04 Std. Gas	0.083	18:43
05 Room Air	0.000	18:44
06 Std. Gas	0.082	18:44
07 Room Air	0.000	18:45

Lot No = 13518080A6  
Cyl No = 20  
Exp Date = 08/05/2020  
County = 49      Oper No. = 058405

  
\_\_\_\_\_  
Operator Signature  
RAY WEBER

Remarks:

Form 106-I8000