



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Dana Rustebakke</i>		
Location <i>BCCC</i>	Intoxilyzer® Serial Number <i>80-006682</i>	
Gas Lot Number <i>13518080A6</i>	Gas Cylinder Number <i>32</i>	Gas Expiration Date <i>8/05/2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C). *ACA TEST DONE*
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.083* AC
 B. *0.083* AC
 C. *0.084* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

*CEE
3.8.19*

Chemical Test Operator Signature <i>Dana Rustebakke</i>	Date <i>3/6/19</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Eden</i>	Date <i>3.8.2019</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006682
Location = BCCC 8164.14.00 09/16
03/06/2019 11:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:22
02 Std. Gas	0.083	11:23
03 Room Air	0.000	11:23
04 Std. Gas	0.083	11:24
05 Room Air	0.000	11:24
06 Std. Gas	0.084	11:24
07 Room Air	0.000	11:25

Lot No = 13518008A6
Cyl No = 32
Exp Date = 08/05/2020
County = 02 Oper No. = 100012



Operator Signature
DANA RUSTEBAKKE

Remarks:

- New Gas Cylinder

Form 106-I8000

