



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL



CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <u>Travis Skar</u>		
Location <u>Burke SO</u>	Intoxilyzer® Serial Number <u>80-006673</u>	
Gas Lot Number <u>13518080AG</u>	Gas Cylinder Number <u>30</u>	Gas Expiration Date <u>08/05/2020</u>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.081 AC
 - B. 0.082 AC
 - C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date <u>3/21/19</u>
Reviewed By (Crime Laboratory Use Only) 	Date <u>4.8.2019</u>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
03/21/2019 09:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:48
02 Std. Gas	0.081	09:48
03 Room Air	0.000	09:49
04 Std. Gas	0.082	09:49
05 Room Air	0.000	09:50
06 Std. Gas	0.082	09:50
07 Room Air	0.000	09:50

Lot No = 13518080A6
Cyl No = 30
Exp Date = 08/05/2020
County = 07 Oper No. = 120513



Operator Signature
TRAVIS SKAR

Remarks:

Form 106-I8000