



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Barry Vannatta</i>		
Location <i>STEE</i>	Intoxilyzer® Serial Number <i>80-006688</i>	
Gas Lot Number <i>34418080A2</i>	Gas Cylinder Number <i>60</i>	Gas Expiration Date <i>2-5-21</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.083 AC
  - B. 0.083 AC
  - C. 0.083 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Barry Vannatta</i>	Date <i>3-25-19</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ed</i>	Date <i>4-8-2019</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = STEE      8164.14.00 09/16  
03/25/2019      14:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:42
02 Std. Gas	0.083	14:42
03 Room Air	0.000	14:43
04 Std. Gas	0.083	14:43
05 Room Air	0.000	14:44
06 Std. Gas	0.083	14:44
07 Room Air	0.000	14:45

Lot No = 34418080A2  
Cyl No = 60  
Exp Date = 02/05/2021  
County = 22      Oper No. = 039909

  
Operator Signature  
BARRY VANNATTA

Remarks:

Form 106-I8000