

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:31
02 Std. Sol.	0.019	08:31
03 Room Air	0.000	08:32
04 Std. Sol.	0.019	08:33
05 Room Air	0.000	08:33
06 Std. Sol.	0.019	08:34
07 Room Air	0.000	08:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 17

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:35
02 Std. Sol.	0.297	08:36
03 Room Air	0.000	08:37
04 Std. Sol.	0.297	08:37
05 Room Air	0.000	08:38
06 Std. Sol.	0.299	08:39
07 Room Air	0.000	08:39


08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 17

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:19

Test	AC	Time
01 Room Air	0.000	08:20
02 *Subject Test	INT*	08:20
03 Room Air	0.000	08:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 8
Expiration Date = 02/05/2020
County = 17 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:22

Test	AC	Time
01 Diagnostic	OK	08:23
02 Room Air	0.000	08:23
03 *Subject Test	RFI*	08:23
04 Room Air	0.000	08:24

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 8
Expiration Date = 02/05/2020
County = 17 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:16
02 Std. Gas	0.080	08:16
03 Room Air	0.000	08:17
04 Std. Gas	0.080	08:17
05 Room Air	0.000	08:18
06 Std. Gas	0.081	08:18
07 Room Air	0.000	08:19

Lot No = 34917080A3
Cyl No = 8
Exp Date = 02/05/2020
County = 17 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:40
02 Std. Gas	0.081	08:41
03 Room Air	0.000	08:41
04 Std. Gas	0.081	08:42
05 Room Air	0.000	08:42
06 Std. Gas	0.081	08:43
07 Room Air	0.000	08:43

Lot No = 34917080A3
Cyl No = 8
Exp Date = 02/05/2020
County = 17 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006667
Location = GLDN 8164.14.00 09/16
04/12/2019 08:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:44
02 Std. Gas	0.081	08:44
03 Room Air	0.000	08:45
04 Std. Gas	0.081	08:45
05 Room Air	0.000	08:45
06 Std. Gas	0.081	08:46
07 Room Air	0.000	08:46

Lot No = 34917080A3
Cyl No = 8
Exp Date = 02/05/2020
County = 17

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000