

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006673

Location: BURKE COUNTY
S.O. (BURK)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 957 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6-8-19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA

Display reads "Interferent Detect"

3. RFI Check (CMS Mode) Display reads "RFI Detect"

4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC

Gas Cyl Lot #: 13518080A6 Cyl #: 30 Exp. Date: 8-5-20

Test 1 0.082 Test 4 0.082 Test 7 0.082

Test 2 0.082 Test 5 0.081 Test 8 0.082

Test 3 0.082 Test 6 0.082 Test 9 0.082

Average 0.082

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eder
Breath Analyst Signature

5-8-2019
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:30

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13518080A6
Standard Cyl #? 30
Standard Expiration? 08/05/2020
Oper No? 666666

Flow Cal. Date: 03/30/2017
Slope 697
Intercept -667725

IR Calibration Date: 03/30/2017
 3um 9um

0th Coef(*100): -9136 -17613
1st Coef(*100): 267126 133333
2nd Coef(*100): 5438 1698
H2O adj(mg/l*10k): 684 589

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Sol.	0.019	12:46
03 Room Air	0.000	12:46
04 Std. Sol.	0.018	12:47
05 Room Air	0.000	12:48
06 Std. Sol.	0.019	12:48
07 Room Air	0.000	12:49

08 Sim Temp = 34.0°C

Simul Ser No = DR3375
Std Sol No = 17160
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:51
02 Std. Sol.	0.295	12:52
03 Room Air	0.000	12:52
04 Std. Sol.	0.297	12:53
05 Room Air	0.000	12:54
06 Std. Sol.	0.297	12:54
07 Room Air	0.000	12:55


08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 07

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:38

Test	AC	Time
01 Room Air	0.000	12:39
02 *Subject Test	INT*	12:39
03 Room Air	0.000	12:40

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 30
Expiration Date = 08/05/2020
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:35

Test	AC	Time
01 Diagnostic	OK	12:36
02 Room Air	0.000	12:36
03 *Subject Test	RFI*	12:37
04 Room Air	0.000	12:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 30
Expiration Date = 08/05/2020
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:32
02 Std. Gas	0.082	12:33
03 Room Air	0.000	12:33
04 Std. Gas	0.082	12:33
05 Room Air	0.000	12:34
06 Std. Gas	0.082	12:34
07 Room Air	0.000	12:35

Lot No = 13518080A6
Cyl No = 30
Exp Date = 08/05/2020
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Gas	0.082	12:41
03 Room Air	0.000	12:41
04 Std. Gas	0.081	12:41
05 Room Air	0.000	12:42
06 Std. Gas	0.082	12:42
07 Room Air	0.000	12:43

Lot No = 13518080A6
Cyl No = 30
Exp Date = 08/05/2020
County = 07

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BURK 8164.14.00 09/16
05/08/2019 12:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Gas	0.082	12:57
03 Room Air	0.000	12:58
04 Std. Gas	0.082	12:58
05 Room Air	0.000	12:59
06 Std. Gas	0.082	12:59
07 Room Air	0.000	12:59

Lot No = 13518080A6
Cyl No = 30
Exp Date = 08/05/2020
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **CALIBRATION CHECK**
0.080 AC

Form 106-I8000