



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>RAY WEBER</i>		
Location <i>Hill</i>	Intoxilyzer® Serial Number <i>80-006678</i>	
Gas Lot Number <i>34418080A2</i>	Gas Cylinder Number <i>58</i>	Gas Expiration Date <i>02-05-2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *081* AC
 B. *081* AC
 C. *081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Ray Weber</i>	Date <i>5-11-2019</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Edr</i>	Date <i>5.20.2019</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006678
Location = HILL 8164.14.00 09/16
05/11/2019 19:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:35
02 Std. Gas	0.081	19:35
03 Room Air	0.000	19:36
04 Std. Gas	0.081	19:36
05 Room Air	0.000	19:37
06 Std. Gas	0.081	19:37
07 Room Air	0.000	19:37

Lot No = 34418080A2
Cyl No = 58
Exp Date = 02/05/2021
County = 49 Oper No. = 058405



Operator Signature
RAY WEBER

Remarks:

Form 106-I8000