

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006507

Location: GRAND FORKS
COUNTY CORRECTIONAL
CENTER
(GFSO)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 708 psi Regulator: 700 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6.8.19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080A2 Cyl #: 55 Exp. Date: 2.5.21

Test 1	<u>0.081</u>	Test 4	<u>0.079</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.081</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: REPLACED BREATH HOSE O-RING,
ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eda
Breath Analyst Signature

5.13.19
Date

NA

NA

Reviewed by _____

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:30

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 55
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 08/20/2015
Slope 659
Intercept -658315

IR Calibration Date: 08/20/2015
 3um 9um

0th Coef(*100): -9452 -11722
1st Coef(*100): 258808 139588
2nd Coef(*100): 4476 1674
H2O adj(mg/l*10k): 499 376

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:45
02 Std. Sol.	0.020	18:45
03 Room Air	0.000	18:46
04 Std. Sol.	0.020	18:47
05 Room Air	0.000	18:47
06 Std. Sol.	0.019	18:48
07 Room Air	0.000	18:48


08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 18

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*Low AC
0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:51
02 Std. Sol.	0.308	18:51
03 Room Air	0.000	18:52
04 Std. Sol.	0.306	18:53
05 Room Air	0.000	18:53
06 Std. Sol.	0.306	18:54
07 Room Air	0.000	18:54

08 Sim Temp = 34.0°C

Simul Ser No = DR7111
Std Sol No = 17350
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:37

Test	AC	Time
01 Room Air	0.000	18:38
02 *Subject Test	INT*	18:38
03 Room Air	0.000	18:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34418080A2
Cyl No = 55
Expiration Date = 02/05/2021
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:35

Test	AC	Time
01 Diagnostic	OK	18:36
02 Room Air	0.000	18:36
03 *Subject Test	RFI*	18:36
04 Room Air	0.000	18:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34418080A2
Cyl No = 55
Expiration Date = 02/05/2021
County = 18 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:32
02 Std. Gas	0.081	18:32
03 Room Air	0.000	18:33
04 Std. Gas	0.080	18:33
05 Room Air	0.000	18:34
06 Std. Gas	0.081	18:34
07 Room Air	0.000	18:35

Lot No = 34418080A2
Cyl No = 55
Exp Date = 02/05/2021
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080A

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:40
02 Std. Gas	0.079	18:40
03 Room Air	0.000	18:41
04 Std. Gas	0.080	18:41
05 Room Air	0.000	18:41
06 Std. Gas	0.080	18:42
07 Room Air	0.000	18:42

Lot No = 34418080A2
Cyl No = 55
Exp Date = 02/05/2021
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC
Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006507
Location = GFSO 8164.14.00 09/16
05/13/2019 18:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:56
02 Std. Gas	0.080	18:56
03 Room Air	0.000	18:57
04 Std. Gas	0.080	18:57
05 Room Air	0.000	18:58
06 Std. Gas	0.080	18:58
07 Room Air	0.000	18:59

Lot No = 34418080A2
Cyl No = 55
Exp Date = 02/05/2021
County = 18 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000