

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007086

Location: LISBON PD  
(LISB)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 966 psi Regulator: 1000 psi  
997 cse

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR5131 Lot #: 201810D Exp. Date: 10.24.20
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 17618080A3 Cyl #: 66 Exp. Date: 8.5.20

Test 1	<u>0.081</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION  
REPLACED EXTERNAL LASER PRINTER

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

5.15.2019  
Date

NA  
Reviewed by

NA  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-007086  
Location = LISB                    8164.16.00 09/18  
05/15/2019                            17:24

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$^&\* () \_ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%\$^&\* () \_ + ?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            17618080A3  
Standard Cyl #?            66  
Standard Expiration?       08/05/2020  
Oper No?                    666666

Flow Cal. Date:            10/16/2018  
Slope                        688  
Intercept                   -653049

IR Calibration Date:       10/16/2018  
                                  3um                            9um

-----  
0th Coef(\*100):        -47466                    -23263  
1st Coef(\*100):        233222                    130299  
2nd Coef(\*100):        3639                      1618  
H2O adj(mg/l\*10k):    1280                      608

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007086  
Location = LISB      8164.16.00 09/18  
05/15/2019      17:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:30
02 Std. Sol.	0.018	17:31
03 Room Air	0.000	17:31
04 Std. Sol.	0.018	17:32
05 Room Air	0.000	17:33
06 Std. Sol.	0.018	17:33
07 Room Air	0.000	17:34


08 Sim Temp = 34.0°C

Simul Ser No = DR5131

Std Sol No = 201810D

County = 37

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*Low AC  
0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007086  
Location = LISB      8164.16.00 09/18  
05/15/2019      17:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:37
02 Std. Sol.	0.306	17:38
03 Room Air	0.000	17:38
04 Std. Sol.	0.304	17:39
05 Room Air	0.000	17:40
06 Std. Sol.	0.304	17:40
07 Room Air	0.000	17:41

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 37

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-007086  
Location = LISB                      8164.16.00 09/18  
05/15/2019                              17:17

Test	AC	Time
01 Room Air	0.000	17:19
02 *Subject Test	INT*	17:19
03 Room Air	0.000	17:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17618080A3  
Cyl No = 66  
Expiration Date = 08/05/2020  
County = 37                              Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check  
0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-007086  
Location = LISB                    8164.16.00 09/18  
05/15/2019                            17:21

Test	AC	Time
01 Diagnostic	OK	17:22
02 Room Air	0.000	17:23
03 *Subject Test	RFI*	17:23
04 Room Air	0.000	17:23

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 17618080A3  
Cyl No = 66  
Expiration Date = 08/05/2020  
County = 37                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007086  
Location = LISB      8164.16.00 09/18  
05/15/2019      17:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:13
02 Std. Gas	0.081	17:13
03 Room Air	0.000	17:14
04 Std. Gas	0.080	17:14
05 Room Air	0.000	17:15
06 Std. Gas	0.080	17:15
07 Room Air	0.000	17:16

Lot No = 17618080A3  
Cyl No = 66  
Exp Date = 08/05/2020  
County = 37      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007086  
Location = LISB      8164.16.00 09/18  
05/15/2019      17:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:25
02 Std. Gas	0.080	17:26
03 Room Air	0.000	17:26
04 Std. Gas	0.080	17:27
05 Room Air	0.000	17:27
06 Std. Gas	0.080	17:28
07 Room Air	0.000	17:28

Lot No = 17618080A3  
Cyl No = 66  
Exp Date = 08/05/2020  
County = 37

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007086  
Location = LISB      8164.16.00 09/18  
05/15/2019      17:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:48
02 Std. Gas	0.080	17:48
03 Room Air	0.000	17:49
04 Std. Gas	0.080	17:49
05 Room Air	0.000	17:50
06 Std. Gas	0.080	17:50
07 Room Air	0.000	17:51

Lot No = 17618080A3  
Cyl No = 66  
Exp Date = 08/05/2020  
County = 37      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000