

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006682

Location: BARNES COUNTY  
CORRECTIONS CENTER  
(BCCC)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 1003 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR5131 Lot #: 201810 D Exp. Date: 10.24.20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 13518080A6 Cyl #: 32 Exp. Date: 8.5.20

Test 1	<u>0.080</u>	Test 4	<u>0.079</u>	Test 7	<u>0.081</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.079</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

UPDATED GAS CYLINDER LOT #:

13518008A6 TO 13518080A6

Instrument is acceptable to be used in the field.

Charles E. Ed  
Breath Analyst Signature

5.15.2019  
Date

NA  
Reviewed by

NA  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006682  
Location = BCCC      8164.14.00 09/16  
05/15/2019      20:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:13
02 Std. Sol.	0.019	20:14
03 Room Air	0.000	20:14
04 Std. Sol.	0.019	20:15
05 Room Air	0.000	20:15
06 Std. Sol.	0.019	20:16
07 Room Air	0.000	20:17

08 Sim Temp = 34.0°C

Simul Ser No = DR5131  
Std Sol No = 201810D  
County = 02      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *Low AC*  
                    *0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006682  
Location = BCCC      8164.14.00 09/16  
05/15/2019      20:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:19
02 Std. Sol.	0.298	20:20
03 Room Air	0.000	20:20
04 Std. Sol.	0.298	20:21
05 Room Air	0.000	20:22
06 Std. Sol.	0.299	20:22
07 Room Air	0.000	20:23


08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 02

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006682  
Location = BCCC      8164.14.00 09/16  
05/15/2019      20:26

Test	AC	Time
01 Room Air	0.000	20:27
02 *Subject Test	INT*	20:27
03 Room Air	0.000	20:28

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 13518080A6  
Cyl No = 32  
Expiration Date = 08/05/2020  
County = 02      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006682  
Location = BCCC                    8164.14.00 09/16  
05/15/2019                            20:28

Test	AC	Time
01 Diagnostic	OK	20:29
02 Room Air	0.000	20:30
03 *Subject Test	RFI*	20:30
04 Room Air	0.000	20:30

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 13518080A6  
Cyl No = 32  
Expiration Date = 08/05/2020  
County = 02                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: **RFI CHECK**

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006682  
Location = BCCC      8164.14.00 09/16  
05/15/2019      20:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:31
02 Std. Gas	0.080	20:31
03 Room Air	0.000	20:32
04 Std. Gas	0.079	20:32
05 Room Air	0.000	20:33
06 Std. Gas	0.080	20:33
07 Room Air	0.000	20:34

Lot No = 13518080A6  
Cyl No = 32  
Exp Date = 08/05/2020  
County = 02      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006682  
Location = BCCC      8164.14.00 09/16  
05/15/2019      20:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:34
02 Std. Gas	0.079	20:35
03 Room Air	0.000	20:35
04 Std. Gas	0.080	20:36
05 Room Air	0.000	20:36
06 Std. Gas	0.080	20:36
07 Room Air	0.000	20:37

Lot No = 13518080A6  
Cyl No = 32  
Exp Date = 08/05/2020  
County = 02      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006682  
Location = BCCC      8164.14.00 09/16  
05/15/2019      20:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:38
02 Std. Gas	0.081	20:38
03 Room Air	0.000	20:39
04 Std. Gas	0.080	20:39
05 Room Air	0.000	20:40
06 Std. Gas	0.079	20:40
07 Room Air	0.000	20:40

Lot No = 13518080A6  
Cyl No = 32  
Exp Date = 08/05/2020  
County = 02      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000