

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006676

Location: NEW SALEM  
(NS) MORTON COUNTY  
S.O.

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 613 psi Regulator: 600 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
  - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR5131 Lot #: 201810 D Exp. Date: 10.24.20
  - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 0.300

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34917080A3 Cyl #: 35 Exp. Date: 2/5/20

Test 1	<u>0.081</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.079</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: REPLACED EXTERNAL LASER PRINTER,  
ANNUAL INSPECTION,  
REPLACED BREATH HOSE O-RING

Instrument is acceptable to be used in the field.

Charles E. Edr  
Breath Analyst Signature

NA

5.23.2019  
Date

NA

Reviewed by

Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = NS      8164.14.00 09/16  
05/23/2019      10:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Sol.	0.018	10:18
03 Room Air	0.000	10:19
04 Std. Sol.	0.018	10:20
05 Room Air	0.000	10:20
06 Std. Sol.	0.017	10:21
07 Room Air	0.000	10:21

08 Sim Temp = 34.0°C

Simul Ser No = DR5131

Std Sol No = 201810D

County = 30

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = NS      8164.14.00 09/16  
05/23/2019      10:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:24
02 Std. Sol.	0.295	10:25
03 Room Air	0.000	10:25
04 Std. Sol.	0.297	10:26
05 Room Air	0.000	10:27
06 Std. Sol.	0.298	10:27
07 Room Air	0.000	10:28

08 Sim Temp = 34.0°C

Simul Ser No = DR7111  
Std Sol No = 17350  
County = 30      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = NS      8164.14.00 09/16  
05/23/2019      09:58

Test	AC	Time
01 Room Air	0.000	09:59
02 *Subject Test	INT*	10:00
03 Room Air	0.000	10:01

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34917080A3  
Cyl No = 35  
Expiration Date = 02/05/2020  
County = 30      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = NS      8164.14.00 09/16  
05/23/2019      09:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:50
02 Std. Gas	0.081	09:50
03 Room Air	0.000	09:50
04 Std. Gas	0.080	09:51
05 Room Air	0.000	09:51
06 Std. Gas	0.080	09:52
07 Room Air	0.000	09:52

Lot No = 34917080A3  
Cyl No = 35  
Exp Date = 02/05/2020  
County = 30

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = NS      8164.14.00 09/16  
05/23/2019      10:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:05
02 Std. Gas	0.080	10:05
03 Room Air	0.000	10:06
04 Std. Gas	0.079	10:06
05 Room Air	0.000	10:06
06 Std. Gas	0.080	10:07
07 Room Air	0.000	10:07

Lot No = 34917080A3  
Cyl No = 35  
Exp Date = 02/05/2020  
County = 30      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = NS      8164.14.00 09/16  
05/23/2019      10:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.081	10:30
03 Room Air	0.000	10:31
04 Std. Gas	0.080	10:31
05 Room Air	0.000	10:32
06 Std. Gas	0.080	10:32
07 Room Air	0.000	10:33

Lot No = 34917080A3  
Cyl No = 35  
Exp Date = 02/05/2020  
County = 30      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000