

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006691

Location: KENMARE P.D.  
(KMPD)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 886 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
    - a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR5131 Lot #: 201810D Exp. Date: 10.24.20
    - b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR3375 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 34418080A2 Cyl #: 49 Exp. Date: 2/5/2021

Test 1	<u>0.081</u>	Test 4	<u>0.081</u>	Test 7	<u>0.080</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.081</u>	Test 6	<u>0.080</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION  
(PERFORMING AT LOCATION EVEN THOUGH ANNUAL INSPECTION  
HAS ALREADY BEEN PERFORMED.)

Instrument is acceptable to be used in the field.

Charles E. Eder  
Breath Analyst Signature NA

5/30/19  
Date NA

Reviewed by \_\_\_\_\_

Date \_\_\_\_\_



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      12:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:13
02 Std. Sol.	0.017	12:14
03 Room Air	0.000	12:15
04 Std. Sol.	0.017	12:15
05 Room Air	0.000	12:16
06 Std. Sol.	0.017	12:17
07 Room Air	0.000	12:17


08 Sim Temp = 34.0°C

Simul Ser No = DR5131

Std Sol No = 201810D

County = 51

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*Low AC*  
*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      12:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:19
02 Std. Sol.	0.293	12:19
03 Room Air	0.000	12:20
04 Std. Sol.	0.293	12:21
05 Room Air	0.000	12:21
06 Std. Sol.	0.295	12:22
07 Room Air	0.000	12:23

08 Sim Temp = 34.0°C

Simul Ser No = DR3375  
Std Sol No = 17350  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: HIGH AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      12:04

Test	AC	Time
01 Room Air	0.000	12:05
02 *Subject Test	INT*	12:06
03 Room Air	0.000	12:07

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 34418080A2  
Cyl No = 49  
Expiration Date = 02/05/2021  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      12:07

Test	AC	Time
01 Diagnostic	OK	12:08
02 Room Air	0.000	12:09
03 *Subject Test	RFI*	12:09
04 Room Air	0.000	12:09

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 34418080A2  
Cyl No = 49  
Expiration Date = 02/05/2021  
County = 51      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      11:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:00
02 Std. Gas	0.081	12:00
03 Room Air	0.000	12:01
04 Std. Gas	0.081	12:01
05 Room Air	0.000	12:02
06 Std. Gas	0.081	12:02
07 Room Air	0.000	12:02

Lot No = 34418080A2  
Cyl No = 49  
Exp Date = 02/05/2021  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      12:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:24
02 Std. Gas	0.081	12:25
03 Room Air	0.000	12:25
04 Std. Gas	0.080	12:26
05 Room Air	0.000	12:26
06 Std. Gas	0.080	12:26
07 Room Air	0.000	12:27

Lot No = 34418080A2  
Cyl No = 49  
Exp Date = 02/05/2021  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006691  
Location = KMPD      8164.14.00 09/16  
05/30/2019      12:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:28
02 Std. Gas	0.080	12:28
03 Room Air	0.000	12:29
04 Std. Gas	0.080	12:29
05 Room Air	0.000	12:30
06 Std. Gas	0.081	12:30
07 Room Air	0.000	12:31

Lot No = 34418080A2  
Cyl No = 49  
Exp Date = 02/05/2021  
County = 51      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080AC*

Form 106-I8000