

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006509 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 452 psi Regulator: 475 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12-12-19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.250 AC
Sim. Ser #: DR7351 Lot #: 201803G Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080A2 Cyl #: 45 Exp. Date: 2-5-21

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.079</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

6-12-19
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 09:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:53
02 Std. Sol.	0.020	09:54
03 Room Air	0.000	09:54
04 Std. Sol.	0.019	09:55
05 Room Air	0.000	09:56
06 Std. Sol.	0.020	09:56
07 Room Air	0.000	09:57

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 201712B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 09:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:00
02 Std. Sol.	0.246	10:01
03 Room Air	0.000	10:01
04 Std. Sol.	0.247	10:02
05 Room Air	0.000	10:02
06 Std. Sol.	0.246	10:03
07 Room Air	0.000	10:04

08 Sim Temp = 34.0°C

Simul Ser No = DR7351

Std Sol No = 201803G

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 10:26

Test	AC	Time
01 Room Air	0.000	10:27
02 *Subject Test	INT*	10:28
03 Room Air	0.000	10:28

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 10:29

Test	AC	Time
01 Diagnostic	OK	10:31
02 Room Air	0.000	10:31
03 *Subject Test	RFI*	10:31
04 Room Air	0.000	10:32

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 10:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.079	10:23
03 Room Air	0.000	10:24
04 Std. Gas	0.080	10:24
05 Room Air	0.000	10:25
06 Std. Gas	0.080	10:25
07 Room Air	0.000	10:26

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 10:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Gas	0.080	10:33
03 Room Air	0.000	10:34
04 Std. Gas	0.079	10:34
05 Room Air	0.000	10:35
06 Std. Gas	0.080	10:35
07 Room Air	0.000	10:36

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
06/12/2019 10:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:37
02 Std. Gas	0.080	10:37
03 Room Air	0.000	10:38
04 Std. Gas	0.080	10:38
05 Room Air	0.000	10:39
06 Std. Gas	0.080	10:39
07 Room Air	0.000	10:39

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000