



# ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Matt O'Brien</i>		
Location <i>Oakes Police Department</i>	Intoxilyzer® Serial Number <i>80-006683</i>	
Gas Lot Number <i>344/8080AZ</i>	Gas Cylinder Number <i>50</i>	Gas Expiration Date <i>02/05/2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
A. 0.078 AC  
B. 0.079 AC  
C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Matt O'Brien</i>	Date <i>9-4-19</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Eda</i>	Date <i>9-5-19</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006683  
Location = OKPD      8164.14.00 09/16  
09/04/2019      17:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:11
02 Std. Gas	0.078	17:11
03 Room Air	0.000	17:12
04 Std. Gas	0.079	17:12
05 Room Air	0.000	17:12
06 Std. Gas	0.079	17:13
07 Room Air	0.000	17:13

Lot No = 34418080A2  
Cyl No = 50  
Exp Date = 02/05/2021  
County = 11      Oper No. = 130690

  
\_\_\_\_\_  
Operator Signature  
MATTHEW O'BRIEN

Remarks:

Form 106-I8000