



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) MARK PEARSON		
Location EMMONS COUNTY	Intoxilyzer® Serial Number 80-000089	
Gas Lot Number 344180 DAZ	Gas Cylinder Number 2	Gas Expiration Date 2-5-2021

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.082 AC
 B. 0.081 AC
 C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

CEE
9.5.19
Rec'd

Chemical Test Operator Signature	Date 9/5/19
Reviewed By (Crime Laboratory Use Only)	Date 9.5.19

Intoxilyzer Test Record and Checklist
 HCOAG Crime Lab. Div., Bismarck, ND 58501
 OVI, Inc. Intoxilyzer Electronic Analyzer
 North Dakota Model 8000 SV 80-000089
 Location = JINT 6164, 14, 00 09/16 10:13
 09/05/2019

DRY CAL CHECK

Test	AC	Time
01 ROOM AIR	0.000	10:13
02 Std. Gas	0.082	10:14
03 ROOM AIR	0.000	10:14
04 Std. Gas	0.081	10:14
05 ROOM AIR	0.000	10:15
06 Std. Gas	0.082	10:15
07 ROOM AIR	0.000	10:15

Lot No = 3441800042
 Cyl No = 2
 Exp Date = 02/05/2021
 County = 15

Operator Signature
 MARK PEARSON

Remarks:

Form 106-18000