



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Jason A. Bullis</i>		
Location <i>Bowman County Sheriff</i>	Intoxilyzer® Serial Number <i>80-006513</i>	
Gas Lot Number <i>34418080A2</i>	Gas Cylinder Number <i>41</i>	Gas Expiration Date <i>2-05-2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.082* AC
 B. *0.082* AC
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>J. Bullis</i>	Date <i>1/24/20</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ehr</i>	Date <i>1/27/20</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006513
Location = BOWM 8164.13.00 06/09
01/24/2020 10:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Gas	0.082	10:03
03 Room Air	0.000	10:04
04 Std. Gas	0.082	10:04
05 Room Air	0.000	10:05
06 Std. Gas	0.082	10:05
07 Room Air	0.000	10:06

Lot No = 34418080A2
Cyl No = 41
Exp Date = 02/05/2021
County = 06 Oper No. = 010703



Operator Signature
JASON BULLIS

Remarks:

Form 106-I8000