



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Kyle Johnson</i>		
Location <i>800 4th Ave E.</i>	Intoxilyzer® Serial Number <i>80-006684</i>	
Gas Lot Number <i>34418080A2</i>	Gas Cylinder Number <i>046</i>	Gas Expiration Date <i>2/5/2021</i>

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function B).
2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.082* AC
 B. *0.082* AC
 C. *0.082* AC
3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> #423	Date <i>2/27/2020</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>3/2/2020</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
02/27/2020 16:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Gas	0.082	16:25
03 Room Air	0.000	16:26
04 Std. Gas	0.082	16:26
05 Room Air	0.000	16:27
06 Std. Gas	0.082	16:27
07 Room Air	0.000	16:27

Lot No = 34418080A2
Cyl No = 46
Exp Date = 02/05/2021
County = 09 Oper No. = 130781



Operator Signature
KYLE JOHNSON

Remarks:

Form 106-I8000



West Fargo Police Department

800 4th Ave E, Suite 2 West Fargo, ND 58078 | 701-433-5500 | westfargopolice.com

Heith R. Janke
Chief of Police

FAX COVER SHEET

TO: Breath Alcohol Toxicology

FROM: Kyle Johnson

RE: _____

DATE: 2/27/2020

PAGES: 2

COMMENTS:

Updated Cylinder

The information contained in this facsimile message may be privileged and confidential. It is intended only for the use of the individual or entity to whom it is sent. If you are not the intended recipient of this telefax message and have received this transmittal by mistake, please do not distribute or copy the communication. Instead, please notify the West Fargo Police Department. Thank you for your cooperation.