



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-006508</b>	Instrument Location <b>RICHLAND COUNTY JAIL</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Matthew Anderson</i>	Date <b>5-28-20</b>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Charles E. Edwards</i>	Certified Date <b>6-4-2020</b>
--	-----------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006508  
Location = WAHP                    8164.13.00 06/09  
05/28/2020                            14:36

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    34418080A2  
Standard Cyl #?                    8  
Standard Expiration?                02/05/2021  
Oper No?                            020401

Flow Cal. Date:                    08/20/2015  
Slope                                644  
Intercept                            -568181

IR Calibration Date:                08/20/2015  
    3um                                9um

-----  
0th Coef(\*100):                    -22365                                -18627  
1st Coef(\*100):                    256969                                136095  
2nd Coef(\*100):                    3168                                    1444  
H2O adj(mg/l\*10k):                693                                    453

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
MATTHEW ANDERSON

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006508  
Location = WAHP                8164.13.00 06/09  
05/28/2020                        14:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.080	14:41
03 Room Air	0.000	14:41
04 Std. Gas	0.081	14:42
05 Room Air	0.000	14:42
06 Std. Gas	0.081	14:43
07 Room Air	0.000	14:43

Lot No = 34418080A2  
Cyl No = 8  
Exp Date = 02/05/2021  
County = 39                        Oper No. = 020401



Operator Signature  
MATTHEW ANDERSON

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006508  
Location = WAHP            8164.13.00 06/09  
05/28/2020            14:45

Test	AC	Time
01 Room Air	RFI*	14:46
02 Room Air	0.000	14:47

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI, TEST NA  
Sub DOB = 05/28/2010  
Sub Sex = Male            Weight = 100  
Test = DUI            Cit = TEST  
Dr. Lic. = ND/TEST  
Lot No = 34418080A2  
Cyl No = 8  
Expiration Date = 02/05/2021  
County = 39            Oper No. = 020401



Operator Signature  
MATTHEW ANDERSON

Remarks:

Form 106-I8000