





**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <b>JASON CARTIER</b>		
Location <b>BURKE County S.O.</b>	Intoxilyzer® Serial Number <b>80-00673</b> <b>CEE</b>	
Gas Lot Number <b>05620080A1</b>	Gas Cylinder Number <b>33</b>	Gas Expiration Date <b>04 05 2022</b>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. **0.083** AC  
 B. **0.082** AC  
 C. **0.083** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:  
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).  
 B. ACA Test Record.

Chemical Test Operator Signature  # <b>4113</b>	Date <b>5/29/20</b>
Reviewed By (Crime Laboratory Use Only) 	Date <b>6.4.2020</b>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006673  
Location = BURK                8164.14.00 09/16  
05/29/2020                        19:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:53
02 Std. Gas	0.083	19:54
03 Room Air	0.000	19:54
04 Std. Gas	0.082	19:55
05 Room Air	0.000	19:55
06 Std. Gas	0.083	19:55
07 Room Air	0.000	19:56

Lot No = 05620080A1  
Cyl No = 33  
Exp Date = 04/05/2022  
County = 07                      Oper No. = 120804

 # 4173  
\_\_\_\_\_  
Operator Signature  
JASON CARTIER

Remarks:

Form 106-I8000