



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number <b>80-006673</b>	Instrument Location <b>BURKE COUNTY J.D.</b>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): NONE
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <b>5/29/20</b>
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date <b>6-4-2020</b>
---------------------------	-----------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006673  
Location = BURK                8164.14.00 09/16  
05/29/2020                      19:38

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05620080A1  
Standard Cyl #?            33  
Standard Expiration?       04/05/2022  
Oper No?                    120804

Flow Cal. Date:            03/30/2017  
Slope                      697  
Intercept                  -667725

IR Calibration Date:        03/30/2017

	3um	9um
0th Coef(*100):	-9136	-17613
1st Coef(*100):	267126	133333
2nd Coef(*100):	5438	1698
H2O adj(mg/l*10k):	684	589

\*\*\*\*\* Printer Test End \*\*\*\*\*

 # 4173  
\_\_\_\_\_  
Operator Signature  
JASON CARTIER

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006673  
Location = BURK      8164.14.00 09/16  
05/29/2020      19:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:53
02 Std. Gas	0.083	19:54
03 Room Air	0.000	19:54
04 Std. Gas	0.082	19:55
05 Room Air	0.000	19:55
06 Std. Gas	0.083	19:55
07 Room Air	0.000	19:56

Lot No = 05620080A1  
Cyl No = 33  
Exp Date = 04/05/2022  
County = 07      Oper No. = 120804

 # 4173  
\_\_\_\_\_  
Operator Signature  
JASON CARTIER

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006673  
Location = BURK                8164.14.00 09/16  
05/29/2020                        19:59

Test	AC	Time
01 Room Air	RFI*	20:04
02 Room Air	0.000	20:04

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, INSTALL NA  
Sub DOB = 06/09/1974  
Sub Sex = Unknown                Weight = NA  
Test = OTH                        Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 05620080A1  
Cyl No = 33  
Expiration Date = 04/05/2022  
County = 07                        Oper No. = 120804

 4/73  
\_\_\_\_\_  
Operator Signature  
JASON CARTIER

Remarks:

Form 106-I8000