



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59281 (06/2018)

Serial Number 80-006683	Instrument Location Oakes Police Department
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>M. A. Brown</i>	Date 6/1/2020
---	------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Charles E. Eh</i>	Certified Date 8.17.20
---	---------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006683  
Location = OKPD                8164.14.00 09/16  
06/01/2020                      14:36

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            34418080A2  
Standard Cyl #?            50  
Standard Expiration?       02/05/2021  
Oper No?                    130690

Flow Cal. Date:            09/06/2016  
Slope                      694  
Intercept                  -595721

IR Calibration Date:       09/06/2016  
                              3um                    9um

-----  
0th Coef(\*100):            -8889                    -21482  
1st Coef(\*100):            254010                   129695  
2nd Coef(\*100):            3922                     1447  
H2O adj(mg/l\*10k):        608                      570

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
MATTHEW O'BRIEN

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006683  
Location = OKPD      8164.14.00 09/16  
06/01/2020      14:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:39
02 Std. Gas	0.080	14:40
03 Room Air	0.000	14:40
04 Std. Gas	0.080	14:41
05 Room Air	0.000	14:41
06 Std. Gas	0.080	14:41
07 Room Air	0.000	14:42

Lot No = 34418080A2

Cyl No = 50

Exp Date = 02/05/2021

County = 11

Oper No. = 130690



Operator Signature  
MATTHEW O'BRIEN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006683  
Location = OKPD      8164.14.00 09/16  
06/01/2020      14:44

Test	AC	Time
01 Room Air	0.000	14:46
02 Reference	RFI*	14:46
03 Room Air	0.000	14:47

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI TEST, NA NA  
Sub DOB = 01/27/1991  
Sub Sex = Male      Weight = NA  
Test = OTH      Cit = NA  
Dr. Lic. = ND/333333  
Lot No = 34418080A2  
Cyl No = 50  
Expiration Date = 02/05/2021  
County = 11      Oper No. = 130690



Operator Signature  
MATTHEW O'BRIEN

Remarks:

Form 106-I8000