

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006675 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 891 psi Regulator: 900 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3061 Lot #: 201810D Exp. Date: 10.29.20
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2.20.21

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3451 Lot #: ICS6 Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 05620080A1 Cyl #: 39 Exp. Date: 4.5.22

|         |              |        |              |        |              |
|---------|--------------|--------|--------------|--------|--------------|
| Test 1  | <u>0.080</u> | Test 4 | <u>0.081</u> | Test 7 | <u>0.080</u> |
| Test 2  | <u>0.080</u> | Test 5 | <u>0.080</u> | Test 8 | <u>0.080</u> |
| Test 3  | <u>0.081</u> | Test 6 | <u>0.081</u> | Test 9 | <u>0.080</u> |
| Average | <u>0.080</u> |        |              |        |              |

C. Remarks/Maintenance: ANNUAL INSPECTION  
TANK SENSOR CALIBRATION

Instrument is acceptable to be used in the field.

Charles E. Eda  
Breath Analyst Signature

6.3.20  
Date

NA  
Reviewed by

NA  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = TOXL      8164.14.00 09/16  
06/03/2020      15:13

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 15:14 |
| 02 Std. Sol. | 0.018 | 15:14 |
| 03 Room Air  | 0.000 | 15:15 |
| 04 Std. Sol. | 0.019 | 15:16 |
| 05 Room Air  | 0.000 | 15:16 |
| 06 Std. Sol. | 0.018 | 15:17 |
| 07 Room Air  | 0.000 | 15:17 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3061  
Std Sol No = 201810D  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*LOW AC*  
*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = TOXL      8164.14.00 09/16  
06/03/2020      15:23

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 15:24 |
| 02 Std. Sol. | 0.295 | 15:25 |
| 03 Room Air  | 0.000 | 15:25 |
| 04 Std. Sol. | 0.296 | 15:26 |
| 05 Room Air  | 0.000 | 15:27 |
| 06 Std. Sol. | 0.295 | 15:27 |
| 07 Room Air  | 0.000 | 15:28 |

08 Sim Temp = 34.0°C

Simul Ser No = MP3069  
Std Sol No = 201902D  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *HIGH AC*  
*0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = TOXL      8164.14.00 09/16  
06/03/2020      15:30

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 15:31 |
| 02 *Subject Test | INT*  | 15:32 |
| 03 Room Air      | 0.000 | 15:32 |

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AE + 0.05% ACETONE*

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006675  
Location = TOXL                    8164.14.00 09/16  
06/03/2020                            15:34

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Diagnostic    | OK    | 15:35 |
| 02 Room Air      | 0.000 | 15:35 |
| 03 *Subject Test | RFI*  | 15:35 |
| 04 Room Air      | 0.000 | 15:36 |

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS921456  
Lot No = 05620080A1  
Cyl No = 34  
Expiration Date = 04/05/2022  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = TOXL      8164.14.00 09/16  
06/03/2020      15:38

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:38 |
| 02 Std. Gas | 0.080 | 15:38 |
| 03 Room Air | 0.000 | 15:39 |
| 04 Std. Gas | 0.080 | 15:39 |
| 05 Room Air | 0.000 | 15:40 |
| 06 Std. Gas | 0.081 | 15:40 |
| 07 Room Air | 0.000 | 15:41 |

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*CALIBRATION CHECK*  
*0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = TOXL      8164.14.00 09/16  
06/03/2020      15:42

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:42 |
| 02 Std. Gas | 0.081 | 15:43 |
| 03 Room Air | 0.000 | 15:43 |
| 04 Std. Gas | 0.080 | 15:44 |
| 05 Room Air | 0.000 | 15:44 |
| 06 Std. Gas | 0.081 | 15:44 |
| 07 Room Air | 0.000 | 15:45 |

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = TOXL      8164.14.00 09/16  
06/03/2020      15:46

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 15:46 |
| 02 Std. Gas | 0.080 | 15:47 |
| 03 Room Air | 0.000 | 15:47 |
| 04 Std. Gas | 0.080 | 15:47 |
| 05 Room Air | 0.000 | 15:48 |
| 06 Std. Gas | 0.080 | 15:48 |
| 07 Room Air | 0.000 | 15:49 |

Lot No = 05620080A1  
Cyl No = 34  
Exp Date = 04/05/2022  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000