



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Dave Feland</i>		
Location <i>Steve</i>	Intoxilyzer® Serial Number <i>80-066688</i>	
Gas Lot Number <i>34418686A2</i>	Gas Cylinder Number <i>60</i>	Gas Expiration Date <i>02-05-2021</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.080* AC
 B. *0.680* AC
 C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 B. ACA Test Record.

Chemical Test Operator Signature <i>David Feland</i>	Date <i>07 Jun 20</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ed</i>	Date <i>06-11-2020</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006688
Location = STEE 8164.14.00 09/16
06/07/2020 15:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:36
02 Std. Gas	0.080	15:36
03 Room Air	0.000	15:36
04 Std. Gas	0.080	15:37
05 Room Air	0.000	15:37
06 Std. Gas	0.080	15:38
07 Room Air	0.000	15:38

Lot No = 34418080A2
Cyl No = 60
Exp Date = 02/05/2021
County = 22 Oper No. = 119404


Operator Signature
DAVID FELAND

Remarks:

Form 106-I8000