



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006677	Instrument Location Minot - Ward County Jail
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **NONE**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature Brian Mellorff	Date 6-10-2020
--	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Charles E. Ed	Certified Date 6-10-2020
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = WARD 8164.14.00 09/16
06/10/2020 15:37

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 7
Standard Expiration? 11/05/2021
Oper No? 050710

Flow Cal. Date: 08/24/2016
Slope 696
Intercept -655873

IR Calibration Date: 08/24/2016

	3um	9um
0th Coef(*100):	-12424	-22862
1st Coef(*100):	263434	131353
2nd Coef(*100):	5070	1477
H2O adj(mg/l*10k):	569	590

***** Printer Test End *****



Operator Signature
BRIAN MEHLHOFF

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = WARD 8164.14.00 09/16
06/10/2020 15:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:38
02 Std. Gas	0.081	15:38
03 Room Air	0.000	15:39
04 Std. Gas	0.082	15:39
05 Room Air	0.000	15:40
06 Std. Gas	0.081	15:40
07 Room Air	0.000	15:41

Lot No = 24119080A1
Cyl No = 7
Exp Date = 11/05/2021
County = 51 Oper No. = 050710



Operator Signature
BRIAN MEHLHOFF

Remarks: ACA Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006677
Location = WARD 8164.14.00 09/16
06/10/2020 15:42

Test	AC	Time
01 Room Air	RFI*	15:45
02 Room Air	RFI*	15:45

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, INSTALL NA
Sub DOB = 12/28/1981
Sub Sex = Unknown Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 24119080A1
Cyl No = 7
Expiration Date = 11/05/2021
County = 51 Oper No. = 050710



Operator Signature
BRIAN MEHLHOFF

Remarks: RFI TEST

Form 106-18000