



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (08/2018)

Serial Number 80-006684	Instrument Location West Fargo Police Dept
Reason for Install/Repair	
<input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests: **CEE 8-17-20 RECEIVED TEST RECORDS**
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): **N/A**
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 6-17-20
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>[Signature]</i>	Certified Date 8-17-2020
---	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
06/17/2020 12:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=-|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 46
Standard Expiration? 02/05/2021
Oper No? 110104

Flow Cal. Date: 09/06/2016
Slope 709
Intercept -725441

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100): -16013 -23467
1st Coef(*100): 256431 135731
2nd Coef(*100): 2904 1268
H2O adj(mg/l*10k): 553 514

***** Printer Test End *****


Operator Signature
TROY NIELSEN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
06/17/2020 12:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Gas	0.080	12:45
03 Room Air	0.000	12:46
04 Std. Gas	0.080	12:46
05 Room Air	0.000	12:47
06 Std. Gas	0.080	12:47
07 Room Air	0.000	12:47

Lot No = 34418080A2
Cyl No = 46
Exp Date = 02/05/2021
County = 09 Oper No. = 110104



Operator Signature
TROY NIELSEN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
06/17/2020 12:53

Test	AC	Time
01 Room Air	0.000	12:54
02 Reference	RFI*	12:54
03 Room Air	0.000	12:55

*Invalid Test
Inhibited - RFI

Sub Name = TEST, RFI N
Sub DOB = 01/01/1990
Sub Sex = Male Weight = 150
Test = DUI Cit = 1234
Dr. Lic. = ND/TES901234
Lot No = 34418080A2
Cyl No = 46
Expiration Date = 02/05/2021
County = 09 Oper No. = 110104


Operator Signature
TROY NIELSEN

Remarks:

Form 106-18000



INTOXILYZER® RECORD
 ND OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 SFN 50496 (05/2018)

Intoxilyzer® Serial Number 80-006684		Location West Fargo Police Dept	
Gas Lot Number 3441 9080A2	Gas Cylinder Number 46	Gas Expiration Date 2-5-21	

Test Date	Chemical Test Operator Number	Operator's Name (PRINT Last Name, First, MI)
6-17-20	110104	Nielsen, Troy N FPD Inshell
6-22-2020	031023	Schwieren, Sylvie 685
6-28-20	110815	Mathson, Chris
6-29-20	131485	Rogstad, Dawson
7-11-20	132140	Saari, David 691
7-12-20	132045	Eink, Travis J 687
7-17-20	130926	Carlson, Michael
7-19-20	131416	Achonen, Brock
7-23-20	130514	Olshann, Marc
7-24-20	121019	HASKELL RYAN W422
8-3-20	510608	Johnson, Travis J W648
8-4-20	121019	UNSWELL, RYAN R W422
8-5-20	131811	Iverson, Paul J FGO #495
8-6-20	130735	Hoxson BEN NDHP 335