

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006671
Location = DEVL 8164.14.00 09/16
06/25/2020 10:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.079	10:23
03 Room Air	RFI*	10:24
04 Room Air	0.000	10:24

*Invalid Test
Inhibited - RFI

Lot No = 24119080A1
Cyl No = 27
Exp Date = 11/05/2021
County = 36 Oper No. = 120501


Operator Signature
JEFFREY BAUSKE

Remarks: RFI TEST

Form 106-I8000