

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006675

Agency being assigned instrument: CRIME LAB - BREATH

Date 7/22/20

Transferred from:

Name (print) Tim Bleth

Signature T. Bleth

Agency Bismarck Police Dept.

Transferred to:

Name (print) CHARLES E. EDER

Signature Charles E. Eder

Agency CRIME LAB - BREATH

Reason for transfer/remarks:

- POWER FAILURE → EEPROM ERROR
- BEING REPLACED BY 80-005945

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____