



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Jordan, Timothy A</i>		
Location <i>Lisb</i>	Intoxilyzer® Serial Number <i>80-007086</i>	
Gas Lot Number <i>05620080 A1</i>	Gas Cylinder Number <i>044</i>	Gas Expiration Date <i>4/5/22</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.081* AC  
 B. *0.281* AC  
 C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>9-21-20</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>10-1-20</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007086  
Location = LISB      8164.16.00 09/18  
09/21/2020      13:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:50
02 Std. Gas	0.081	13:50
03 Room Air	0.000	13:51
04 Std. Gas	0.081	13:51
05 Room Air	0.000	13:52
06 Std. Gas	0.081	13:52
07 Room Air	0.000	13:52

Lot No = 05620080A1  
Cyl No = 44  
Exp Date = 04/05/2022  
County = 37      Oper No. = 020607

  
\_\_\_\_\_  
Operator Signature  
TIM JORDAN

Remarks:

Form 106-I8000