



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) JON WOLD		
Location SARGENT COUNTY SO SAGO		Intoxilyzer® Serial Number 80-006681
Gas Lot Number 05620080A1	Gas Cylinder Number 42	Gas Expiration Date 4/05/2022

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C). **REC'D ACA TEST**
- CEE** Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. **0.080** AC
 - B. **0.080** AC
 - C. **0.079** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory: **REC'D BOTH FORMS**
 - CEE** A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature Jon Wold	Date 10/23/2020
Reviewed By (Crime Laboratory Use Only) Charles E. Ed	Date 11/3/2020


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006681
Location = SACO 8164.14.00 09/16
10/23/2020 18:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:29
02 Std. Gas	0.080	18:30
03 Room Air	0.000	18:30
04 Std. Gas	0.080	18:30
05 Room Air	0.000	18:31
06 Std. Gas	0.079	18:31
07 Room Air	0.000	18:32

Lot No = 05620080A1
Cyl No = 42
Exp Date = 04/05/2022
County = 41 Oper No. = 132607



Operator Signature
JONATHAN WOLD

Remarks:

Form 106-I8000