

**CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT**

Instrument: 8000 Serial Number: 80-006669

Agency being assigned instrument: PROPERTY DISPOSAL - DUMP

Date 12/3/2020

Transferred from: Name (print) CHARLES E. EDER Signature *Charles E. Eder* Agency D.A.G. - CRIME LAB

Transferred to: Name (print) PROPERTY DISPOSAL Signature _____ Agency DUMP

Reason for transfer/remarks:
- EEPROM 18 ERROR → NON FUNCTIONAL
- NOT SUITABLE FOR REPAIR DUE TO COST
- ELECTRONICS & LANDFILL DISPOSAL

Or Shipped Via:
USPS _____ UPS _____ FedEx _____ Other _____


Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____



PROPERTY DISPOSAL
OFFICE OF ATTORNEY GENERAL
 SFN 60567 (7/2016)


Multiple quantities of like items may be listed on one form and attach a sheet with tag numbers and serial numbers if available. Use a separate form for different items.

Division Crime Laboratory	Location of Property Crime Lab - Breath Section	Date 12/3/20
Point of Contact Charles E. Eder	Point of Contact Email ceder@nd.gov	Point of Contact Telephone Number 701-328-6159

Description of Property (make, model, color, etc.) Intoxilyzer 8000 Evidential Breath Alcohol Instrument 	Tag Number (if available) AG 3567
	Serial Number (if available) 80-006669
	Date Purchased (if available)
	Original Cost (if available)
Reason for disposal EEPROM 18 Error - Non Functional. Not suitable for repair due to cost.	

Method of disposal:

- Landfill
- ELECTRONICS DISPOSAL**
- Trash can/dumpster

Signature of person disposing property 
Date 12/3/2020

Send form to Michelle Metzger, Finance Division

Authorized Representative Signature
Date